

GX-A SERIES

GF-A SERIES

Multi-Function Balance

INSTRUCTION MANUAL

G X - A series

GX-203A/GX-303A/GX-403A/GX-603A/GX-1003A/GX-1603A
GX-2002A/GX-3002A/GX-4002A/GX-6002A/GX-10002A
GX-6001A/GX-10001A

G F - A series

GF-203A/GF-303A/GF-403A/GF-603A/GF-1003A/GF-1603A
GF-2002A/GF-3002A/GF-4002A/GF-6002A/GF-10002A
GF-6001A/GF-10001A



1WMPD4003475

© 2017 A&D Company Ltd. All rights reserved.

No part of this publication may be reproduced, transmitted, transcribed, or translated into any language in any form by any means without the written permission of A&D Company Ltd.

The contents of this manual and the specifications of the instrument covered by this manual are subject to change for improvement without notice.

Windows, Word and Excel are registered trademarks of the Microsoft Corporation.

Contents

1. Introduction	4
1-1 Features	4
1-2 About The Models	5
1-3 Compliance	5
2. Unpacking The Balance	7
2-1 Installing The Balance	9
2-2 Precautions	9
2-3 During Use	10
2-4 After Use.....	11
2-5 Power Supply.....	11
3. Display Symbols And Key Operation.....	12
4. Weighing Units	14
4-1 Units	14
4-2 Storing Units.....	17
5. Weighing	19
5-1 Basic Operation.....	19
5-2 Counting Mode (PCS)	21
5-3 Percent Mode (%).....	23
5-4 Animal Weighing Mode (Hold Function)	24
6. Impact Detection Function	25
7. Response Adjustment / Self Check Function	26
7-1 Response Adjustment	26
7-2 Self-Check-Function / Automatic Setting Of Minimum Weight Value.....	27
8. Calibration.....	28
8-1 Automatic Self Calibration For The GX-A Series	29
8-2 One-Touch Calibration For The GX-A Series.....	30
8-3 Calibration Using An External Weight	31
8-4 Calibration Test Using An External Weight	32
8-5 Correcting The Internal Mass Value Of The GX-A series	33
8-6 Correcting The Internal Mass Value Of The GX-A series (Auto)	34
8-7 Correcting The Internal Mass Value Of The GX-A series (Manual)	35
9. Function Switch And Initialization	36
9-1 Permit Or Inhibit	36

9-2 Initializing The Balance	38
10. Function Table	39
10-1 Setting The Function Table	39
10-2 Details Of The Function Table.....	41
10-3 Description Of The Class "Environment, Display"	48
10-4 Clock And Calendar Function	50
10-5 Comparator Function	52
11. ID Number And GLP Report.....	55
11-1 Setting The ID Number	55
11-2 GLP Report	57
12. Data Memory	60
12-1 Data Memory For Weighing Data	60
12-2 Data Memory For Calibration And Calibration Test	62
13. Statistical Calculation Mode	65
13-1 How To Use The Statistic Calculation	65
13-2 Statistical Calculation Mode (Example Of Use).....	70
14. Flow Measurement	72
14-1 How To Use Flow Measurement.....	73
14-2 Flow Measurement Setting	75
15. Gross Net Tare Function	77
15-1 Preparation Of Gross Net Tare Function	77
15-2 Example Of Using The Gross Net Tare Function	78
16. Minimum Weighing Warning Function	79
17. Underhook	81
18. Programmable-Unit	82
19. Density Measurement.....	83
20. Password Lock Function.....	88
20-1 Using Password Function	88
20-2 Changing Password.....	89
20-3 Inputting Password When Turning On The Balance.....	91
20-4 Missing Password	91
21. Maintenance	92

21-1 Treatment Of The Balance	92
22. Troubleshooting.....	92
22-1 Checking The Balance Performance And Environment.....	92
22-2 Error Codes	94
22-3 Other Display	97
22-4 Asking For Repair.....	97
23. Specififations	98
23-1 GX-A series 0.001g models.....	98
23-2 GX-A series 0.01g models.....	99
23-3 GX-A series 0.1g models.....	100
23-4 GF-A series 0.001g models.....	101
23-5 GF-A series 0.01g models.....	102
23-6 GF-A series 0.1g models	103
24. External Dimention.....	104
24-1 Options And Perpheral Instruments	106

1. Introduction

This manual describes how the GX-A/GF-A series balance works and how to get the most out of it in terms of performance. Read this manual thoroughly before using the balance and keep it at hand for future reference.

This manual consists of the following five parts:

- Basic operationDescribes precautions on handling the balance, balance construction and basic balance operation.
- Adapting to the environment Describes response adjustment, calibration and calibration test.
- Functions.....Describes various functions of the balance.
- Interface Describes the interface which transmits data and controls the balance. To use it, need a personal computer or an optional printer.
- MaintenanceDescribes maintenance, error codes, troubleshooting, specifications and options.

1-1 Features

- The balance has a self-check function that inspects the balance itself using electronically controlled load (ECL) and evaluates performance. Read this manual thoroughly before using the balance and keep it at hand for future reference.
- The balance can detect impact applied to its mass sensor and display the level of that impact. ISD (Impact Shock Detection).
- Continuous change of the balance can be calculated as flow rate, displayed and output. FRD :(Flow Rate Display).
- The balance is equipped with a data memory function, which can record weighing value, calibration result, and multiple unit mass (mass per sample in counting mode) (Up to 200 items are stored for weighing value).
- The GX-A series has automatic self calibration using the internal mass, adapting to temperature changes, setting time and interval time.
- Good laboratory practice (GLP) / Good manufacturing practice (GMP) data can be output using the RS-232C serial interface.
- A built-in clock and calendar that can add the time and date to the output data.
- Comparator Indicators, displaying the comparison results with . (Depending on the setting, 5-step comparison is also possible.)
- Capacity Indicator, displaying the weight value in percentage relative to the weighing capacity.
- Hold Function, provided for weighing a moving object such as an animal.
- Underhook, for measuring density and weighing magnetic materials.
- Users of the balance can be limited by setting a password (Password lock function).
- The balance is equipped with an RS-232C serial interface and a USB interface to communicate with a computer. Windows computer using the Windows communication tools software (WinCT) make building a system very easy. The latest Win-CT software can be downloaded from the A&D website. Windows is the registered trademark of the Microsoft Corporation.
- A small breeze break is included with the model featuring a minimum display of 0.001g.

1-2 About The Models

There are many models in the GX-A series and GF-A series with differences in the models being the minimum display and weighing capacity. In this manual, they are listed collectively by the minimum display as shown in the table below.

Model	Minimum display	Applicable model	
		Internal mass type	General type
0.001g model	0.001g	GX-203A / GX-303A / GX-403A / GX-603A / GX-1003A / GX-1603A	GF-203A / GF-303A / GF-403A / GF-603A / GF-1003A / GF-1603A
0.01g model	0.01g	GX-2002A / GX-3002A / GX-4002A / GX-6002A / GX-10002A	GF-2002A / GF-3002A / GF-4002A / GF-6002A / GF-10002A
0.1g model	0.1g	GX-6001A / GX-10001A	GF-6001A / GF-10001A

- For the GX-A series, a weight for sensitivity adjustment is built in. It is possible to use functions such as calibration and auto calibration using the internal mass.
- For the GF-A series, sensitivity adjustment weights are not built-in. When calibrating, it is necessary to prepare an external weight.

1-3 Compliance

Compliance with FCC Rules

Please note that this equipment generates, uses and can radiate radio frequency energy. This equipment has been tested and has been found to comply with the limits of a Class A computing device pursuant to Subpart J of Part 15 of FCC rules. These rules are designed to provide reasonable protection against interference when equipment is operated in a commercial environment. If this unit is operated in a residential area, it may cause some interference and under these circumstances the user would be required to take, at his own expense, whatever measures are necessary to eliminate the interference.

(FCC = Federal Communications Commission in the U.S.A.)

Compliance With Directives of CE mark

CE This device features radio interference suppression, safety regulation and restriction of Hazardous Substances in compliance with the following Council Directives

Council directive 2014/30/EU EN61326 EMC directive

Council directive 2014/35/EU EN60950 Safety of Information Technology Equipment

Council directive 2011/65/EU EN50581 Restriction of the use of certain Hazardous Substances

The CE mark is an official mandatory European marking.

Please note that any electronic product must comply with local laws and regulations when sold or used anywhere outside Europe.



A & D Instruments Ltd. hereby declare that the following Weighing product conforms to the requirements of the council directives on ...

**Electromagnetic Compatibility (EMC) 2014/30/EU,
Low Voltage Equipment (LVD) 2014/35/EU and
Restriction of the use of certain Hazardous Substances (RoHS) 2011/65/EU**

provided that they bear the CE mark of conformity.

Model/Series....GX-A/GF-A Series

Standards applicable:

EN 61326-1:2013

Electrical equipment for measurement, control and laboratory use -EMC requirements Part 1: General requirements

EN 60950-1:2006+A11:2009+A1:2010+A12:2011+A2:2013


Safety of Information Technology Equipment

EN 50581:2012

Technical documentation for the assessment of electrical and electronic products with respect to the restriction of hazardous substances

CE Mark first applied 16 October 2017

Signed for A&D Instruments in Oxford England 27 November 2017




J. Ghuman
Managing Director

2. Unpacking The Balance

The balance is a precision instrument. Unpack the balance carefully. Keep the packing material to be used for transporting the balance in the future.


The packing contents depend on the balance model. See the illustrations to confirm that everything is contained. When shipping options are included, optional accessories may be bundled.

GX-A / GF-A 0.001g models



Note : Please confirm that the AC adapter type is correct for your local voltage and receptacle type.

GX-A / GF-A 0.01g/0.1g models



Note : Please confirm that the AC adapter type is correct for your local voltage and receptacle type.

2-1 Installing The Balance


Install the balance as follows:

1. Refer to "2-2. PRECAUTIONS" for installing the balance.
2. Assemble the balance as shown in the illustration above.
3. Adjust the leveling feet to level the balance. Confirm it using the bubble spirit level.
4. Confirm that the adapter type is correct for the local voltage and power receptacle type.
5. Connect the AC adapter to the balance. Warm up the balance for at least 30 minutes with nothing on the weighing pan.

2-2 Precautions

To get the optimum performance from the balance and acquire accurate weighing data, note the following:



- Install the balance in an environment where the temperature and humidity are not excessive. The best operating temperature is about $20^{\circ}\text{C}\pm 2^{\circ}\text{C}$ at about 45~60%RH relative humidity.
- Install the balance where it is free of dust.
- The weighing table should be solid and free from vibration, drafts and as level as possible.
- Install the balance in a stable place avoiding vibration and shock. Corners of rooms on the first floor are best, as they are less prone to vibration.
- Install the balance where it is not affected by heaters or air conditioners.
- Install the balance where it is not exposed to direct sunlight.
- Install the balance away from equipment which produces magnetic fields.
- Level the balance by adjusting the leveling feet and confirm it using the bubble spirit level.
- Warm up the balance for at least 30 minutes. Plug in the AC adapter as usual.
- Calibrate the balance before use or after having moved it to another location. Refer to "8.Calibration".



Caution

Do not install the balance where flammable or corrosive gas is present.



How to adjust the bubble spirit level






2-3 During Use

To acquire accurate weighing data, note the following:

- Discharge static electricity from the material to be weighed. When a sample could have a static charge, the weighing data is influenced. If the ambient humidity becomes 45% or less, insulators such as plastics are liable to become static electricity. Ground the balance and try the following.
 - Eliminate the static electricity by GXA-25, AD-1683 as an accessory.
 - Or try to keep highly the ambient humidity
 - Or use a metal shield case.
 - Or wipe a charged plastic sample with the wet cloth.
- The breeze break (1mg models only) and the clear main unit cover are provided as accessories. The breeze break components may be charged with static electricity when they are unpacked or when the humidity is low. If the weighing value is unstable or the balance has a problem with repeatability, remove the breeze break. Or wipe the clear plates with a moistened cloth, use an accessory DC static eliminator, GXA-25, AD-1683 or apply an anti-static spray.
- This balance uses a strong magnet as part of the balance assembly, so please use caution when weighing magnetic materials such as iron. If there is a problem, use the underhook on the bottom of the balance to suspend the material away from the influence of the magnet.



- Cancel the temperature difference between a sample and the environment. When a sample is warmer (cooler) than the ambient temperature, the sample will be lighter (heavier) than the true weight. This error is due to a rising (falling) draft around the sample.
- Make each weighing gently and quickly to avoid errors due to changes in the environmental conditions.
- When placing a sample on a weighing pan, do not give a strong shock or do not exceed the weighing capacity. And place in the center.
- Do not drop things upon the weighing pan, or place a sample on the pan that is beyond the balance weighing capacity. Place a sample in the center of the weighing pan.
- Do not use a sharp instrument such as a pencil to press the keys. Use your finger only.
- Press the **RE-ZERO** key before each weighing to prevent possible errors.
- Take into consideration the affect of air buoyancy on a sample when more accuracy is required.
- Keep the balance interior free of dust and foreign materials.



2-4 After Use

- Avoid mechanical shock to the balance.
- Do not disassemble the balance. Contact the local A&D dealer if the balance needs service or repair.
- Do not use organic solvents to clean the balance. Clean the balance with a lint free cloth that is moistened with warm water and a mild detergent.
- Avoid dust and water so that the balance weighs correctly. Protect the internal parts from liquid spills and excessive dust.


2-5 Power Supply

- When the AC adapter is connected, the balance is in the standby mode if the standby indicator is on. This is a normal state and does not harm the balance. For accurate weighing, keep the AC adapter connected to the balance and AC power unless the balance is not to be used for a long period of time.


3. Display Symbols And Key Operation

Display symbols

- Number of statistical data (Statistical calculation mode)
- Displays the weight data relative to the weighing capacity, in percentage, in the weighing mode (**Capacity indicator**)



Blinking display contents




Key operation

Key operation affects how the balance functions. The basic key operations are:







- “Press and release the key immediately” or “Press the key”
= normal key operation during measurement
- “Press and hold the key”



Press the key.
(Press and release the key immediately.)




Press and hold the key.

Key	When pressed	When pressed and held
	Turns the display ON:OFF . The standby indicator is displayed when the display is turned off. The weighing mode is enabled when the display is turned on. When password function is enable, password input display will be displayed. Refer to "20-3 Inputting Password Lock When Turning On The Balance" This ON:OFF key is available anytime. Pressing the ON:OFF key during operation will interrupt operation and turn the display OFF.	
	In the weighing mode, turns the minimum weighing value on and off. In the counting or percent mode, enters the sample storing mode.	Enters the function table mode.
	Switches the weighing units stored in the function table. (<i>g, PLS, %, ct, mom</i>)	Displays other items of the calibration menu.
	Performs calibration of the balance using the internal mass.	Displays other items of the calibration menu.
	Stores the weighing data in memory or outputs to a printer or personal computer depending on the function table settings. (Factory setting = output)	Enters mode to change the unit mass registration number in counting mode. By changing the function table: • Outputs "Title block" and "End block" for GLP,GMP report. • Displays the data memory menu. • Enters mode for reading density number in flow measurement.
	Sets the display to zero.	

4. Weighing Units

4-1 Units

With the GX-A /GF-A series balance, the following weighing units and weighing modes are available :



A unit or mode can be selected and stored in the function table as described on page 15.

If a weighing mode (or unit of weight) has been turned off, that mode or unit will be missing in the sequence. Tael has four varieties, one of which can be selected and installed at the factory.

To select a unit or mode for weighing, press the **MODE** key.

For details about the units and modes, see the table below:

Name (unit, mode)	Abbrev.	Display	Function table (Storing mode)	Conversion factor 1 g =
Gram	g	g	g	1 g
Counting mode	PCS	PCS	PCS	—
Percent mode	%	%	%	—
Ounce (Avoir)	OZ	OZ	OZ	28.349523125 g
Pound	Lb	Lb	Lb	453.59237 g
Pound/Ounce	L OZ	L OZ	LO	1Lb=16 oz, 1 oz=28.349523125 g
Troy Ounce	OZt	OZt	OZt	31.1034768 g
Metric Carat	ct	ct	ct	0.2 g
Momme	mom	mom	mom	3.75 g
Pennyweight	dwt	dwt	dwt	1.55517384 g
Grain (UK)	GN	GN	GN	0.06479891 g
Tael (HK general, Singapore)	TL	TL	TL	37.7994 g
Tael (HK jewelry)				37.429 g
Tael (Taiwan)				37.5 g
Tael (China)				31.25 g
Tola (India)	tol	tol	tol	11.6638038 g
Messghal	MES	MES	MES	4.6875 g
Density mode (See note below)	DS	DS is used to show the density.	DS	—
Programmable-unit (Multi-unit)	MLT	MLt	MLt	—

Note: The blinking processing indicator with "g" indicates that the density mode is selected.

The tables below indicate the weighing capacity and the minimum display for each unit, depending on the balance model.

Unit	GX-203A	GX-303A	GX-403A	GX-603A	GX-1003A	GX-1603A	Minimum display
	GF-203A	GF-303A	GF-403A	GF-603A	GF-1003A	GF-1603A	
	Capacity						
Gram	220	320	420	620	1100	1620	0.001
Ounce (Avoir)	7.76	11.28	14.81	21.86	38.80	57.14	0.00005
Pound	0.485	0.705	0.925	1.366	2.425	3.571	0.000005
Pound/Ounce	0Lb 7.76oz	0Lb 11.28oz	0Lb 14.81oz	1Lb 5.86oz	2Lb 6.80oz	3Lb 9.14oz	0.01oz
Troy Ounce	7.07	10.28	13.50	19.93	35.36	52.08	0.00005
Metric Carat	1100	1600	2100	3100	5500	8100	0.005
Momme	58.6	85.3	112.0	165.3	293.3	432.0	0.0005
Pennyweight	141	205	270	398	707	1041	0.001
Grain (UK)	3395	4938	6481	9568	16975	25000	0.02
Tael (HK general, Singapore)	5.82	8.46	11.11	16.40	29.10	42.85	0.00005
Tael (HK jewelry)	5.87	8.54	11.22	16.56	29.38	43.28	0.00005
Tael (Taiwan)	5.86	8.53	11.20	16.53	29.33	43.20	0.00005
Tael (China)	7.04	10.24	13.44	19.84	35.20	51.84	0.00005
Tola (India)	18.8	27.4	36.0	53.1	94.3	138.8	0.0001
Messghal	46.9	68.2	89.6	132.2	234.6	345.6	0.0005

Unit	GX-2002A	GX-3002A	GX-4002A	GX-6002A	GX-10002A	Minimum display
	GF-2002A	GF-3002A	GF-4002A	GF-6002A	GF-10002A	
	Capacity					
Gram	2200	3200	4200	6200	10200	0.01
Ounce (Avoir)	77.6	112.8	148.1	218.6	359.7	0.0005
Pound	4.85	7.05	9.25	13.66	22.48	0.00005
Pound/Ounce	4Lb 13.60oz	7Lb 0.87oz	9Lb 4.15oz	13Lb 10.69oz	22Lb 7.79oz	0.01oz
Troy Ounce	70.7	102.8	135.0	199.3	327.9	0.0005
Metric Carat	11000	16000	21000	31000	51000	0.05
Momme	586	853	1120	1653	2720	0.005
Pennyweight	1414	2057	2700	3986	6558	0.01
Grain (UK)	33951	49383	64815	95680	157410	0.2
Tael (HK general, Singapore)	58.2	84.6	111.1	164.0	269.8	0.0005
Tael (HK jewelry)	58.7	85.4	112.2	165.6	272.5	0.0005
Tael (Taiwan)	58.6	85.3	112.0	165.3	272.0	0.0005
Tael (China)	70.4	102.4	134.4	198.4	326.4	0.0005
Tola (India)	188	274	360	531	874	0.001
Messghal	769	682	896	1322	2176	0.005

Unit	GX-6001A	GX-8001A	Minimum display
	GF-6001A	GF-8001A	
	Capacity		
Gram	6200	10200	0.1
Ounce (Avoir)	218	359	0.005
Pound	13.6	22.4	0.0005
Pound/Ounce	13Lb 10.69oz	22Lb 7.79oz	0.01oz
Troy Ounce	199	327	0.005
Metric Carat	31000	51000	0.5
Momme	1653	2720	0.05
Pennyweight	3986	6558	0.1
Grain (UK)	95680	157410	2
Tael (HK general, Singapore)	164.0	269.0	0.005
Tael (HK jewelry)	165.0	272.0	0.005
Tael (Taiwan)	165.0	272.0	0.005
Tael (China)	198.0	326.0	0.005
Tola (India)	531.0	874.0	0.01
Messghal	1322	2176	0.05

4-2 Storing Units

The units or modes can be selected and stored in the function table. The sequence of displaying the units or modes can be arranged to fit the frequency of use.

The units stored are maintained in non-volatile memory, even if the AC adapter is removed.

Select a unit or mode and arrange the sequence of display as follows:

1 Press and hold the **SAMPLE** key until **bASFnC** of the function table is displayed, then release the key.

2 Press the **SAMPLE** key several times to display **Unit**.

3 Press the **PRINT** key to enter the unit selection mode.

4 Specify a unit or mode in the order to be displayed using the following keys.

SAMPLE key

To sequentially display the units.

RE-ZERO key


To specify a unit or mode. The stabilization indicator **○** appears when the displayed unit or mode is specified.

If the key is pressed in units already selected, the stability mark disappears.

5 Press the **PRINT** key to store the units or modes. The balance displays **End** and then displays the next menu of the function table.

6 Press the **CAL** key to exit the function table. Then the balance returns to the weighing mode with the selected unit.


7 To select other unit or mode for weighing, press the **MODE** key.



Unit setting example

The example below sets the units in the order with g (gram) as the first unit followed by pc (counting mode).

- 1 Press and hold the **SAMPLE** key until **bRSFnC** of the function table is displayed, then release the key.
- 2 Press the **SAMPLE** key several times to display **Unit**.
- 3 Press the **PRINT** key to enter the unit selection mode.
- 4 Press the **RE-ZERO** key to specify the unit of g
The stabilization indicator **○** appears when the unit is specified.
- 5 Press the **SAMPLE** key to display **Unit PCS**.
- 6 Press the **RE-ZERO** key to specify the unit of pc
The stabilization indicator **○** appears when the unit is specified.
- 7 Press the **PRINT** key to store the units.
The balance displays **End** and then displays the next menu item of the function table.
- 8 Press the **CAL** key to exit the function table. Then the balance returns to the weighing mode with g, the unit selected first.
- 9 Press the **MODE** key to switch between g and pc (g→pcs).



5. Weighing

5-1 Basic Operation

1. Press **MODE** key, and then select the appropriate units (**g** , **ct** , **mg**)


In this case, select "g".

2. Place a container on the weighing pan, if necessary.

Press the **RE-ZERO** key to cancel the weight (tare). The balance displays **0.00 g**. (The decimal point position depends on the balance model.)

3. Place a sample on the pan or in the container. Wait for the stabilization indicator **○** to be displayed. Read the value.

Remove the sample and container from the pan.



Notes

- Press the **SAMPLE** key to turn on or off the minimum weighing value.
- The weighing data can be stored in memory. For details, refer to "12. Data Memory".
- When the **ON:OFF** key is pressed with a container placed on the weighing pan and weighing is started, the balance automatically cancels the weight (tare) and displays **0.00 g**.

About the operation at when power is turned on

The balance will decide the reference zero point when the power is turned on (AC adapter is connected). Depending on the load condition at that time, it will automatically judge whether to perform zero or tare operation. The condition for determining which is used is "power on zero range", and when power on zero range is exceeded, the tare subtraction operation is performed.

About re-zero operation

By pressing the **RE-ZERO** key, the display can be changed to zero.

Re-zero with the **RE-ZERO** key will automatically determine whether zero or tare operation is performed.

The condition for determining which is used is "zero range", and when zero range is exceeded, the tare subtraction operation is performed.

About measurement range

For the balance, the range that can be weighed is determined by model.

The total amount (net amount + tare quantity) up to the maximum display of each model is displayed, and when the maximum display is exceeded, **E** is displayed to indicate that the weighing range is exceeded.

When in excess in negative, **-E** is displayed.

Model		Power on zero range	Zero range	-E display range
GX-203A	GF-203A	Approx. $\pm 100\text{g}$	Approx. $\pm 4\text{g}$	Approx. -100g or less
GX-303A	GF-303A	Approx. $\pm 100\text{g}$	Approx. $\pm 6\text{g}$	Approx. -100g or less
GX-403A	GF-403A	Approx. $\pm 100\text{g}$	Approx. $\pm 8\text{g}$	Approx. -100g or less
GX-603A	GF-603A	Approx. $\pm 100\text{g}$	Approx. $\pm 12\text{g}$	Approx. -100g or less
GX-1003A	GF-1003A	Approx. $\pm 100\text{g}$	Approx. $\pm 20\text{g}$	Approx. -100g or less
GX-1603A	GF-1603A	Approx. $\pm 100\text{g}$	Approx. $\pm 32\text{g}$	Approx. -100g or less
GX-2002A	GF-2002A	Approx. $\pm 1\text{kg}$	Approx. $\pm 40\text{g}$	Approx. -1kg or less
GX-3002A	GF-3002A	Approx. $\pm 1\text{kg}$	Approx. $\pm 60\text{g}$	Approx. -1kg or less
GX-4002A	GF-4002A	Approx. $\pm 1\text{kg}$	Approx. $\pm 80\text{g}$	Approx. -1kg or less
GX-6002A	GF-6002A	Approx. $\pm 1\text{kg}$	Approx. $\pm 120\text{g}$	Approx. -1kg or less
GX-10002A	GF-10002A	Approx. $\pm 1\text{kg}$	Approx. $\pm 200\text{g}$	Approx. -1kg or less
GX-6001A	GF-6001A	Approx. $\pm 1\text{kg}$	Approx. $\pm 120\text{g}$	Approx. -1kg or less
GX-10001A	GF-10001A	Approx. $\pm 1\text{kg}$	Approx. $\pm 200\text{g}$	Approx. -1kg or less

5-2 Counting Mode (PCS)

This is the mode to determine the number of objects in a sample based on the standard sample unit mass. Unit mass means the mass of one sample. The smaller the variables in each sample unit mass is, the more accurate the counting will be. This series balance is equipped with the Automatic Counting Accuracy Improvement (ACAI) function to improve the counting accuracy.

Note

- ※ For counting, use samples that have a unit mass at least ten times greater than that of the minimum display in grams.
- ※ If the sample unit mass variable is too large, it may cause a counting error.
- ※ To improve the counting performance, use the ACAI function frequently or divide the samples into several groups and count each group.

Selecting the counting mode

1. Press the **MODE** key to select **PCS** (PCS = unit)

Storing a sample unit mass

2. Press the **SAMPLE** key to enter the sample unit mass storing mode. Even in the storing mode, pressing the **MODE** key will switch to the next mode.
3. To select the number of samples, press the **SAMPLE** key several times. It may be set to 5, 10, 25, 50 or 100.

Note

A greater number of samples will yield more accurate counting result.


4. Place a container on the weighing pan, if necessary. Press the **RE-ZERO** key to cancel the weight (tare). The number specified in step 3 appears. **25 0** is displayed if 25 is selected in step 3.
5. Place the number of samples specified on the pan. In this example, 25 pieces.
6. When **PRINT** key pressed, unit mass is stored and changes the count display. (eg: when the number is 25, **25 PCS** is displayed).

Note

- ※ If the balance judges that the mass of the samples is TOO light to acquire accurate weighing, it displays an error requiring the addition of more samples to the specified number and press the **PRINT** key. When the unit mass is stored correctly, the balance proceeds to the counting mode.
- ※ If the balance judges that the mass of the samples is too light and is not adequate to be used as the unit mass, it displays **Lo**.
- ※ Registered unit mass is remembered even when the power is turned off.

Number mode (counting)

7. Counting is possible.




Counting Mode Using The ACAI Function

The ACAI is a function that improves the accuracy of the unit mass automatically by increasing the number of samples as the counting process.

ACAI: Automatic Counting Accuracy Improvement

After registering unit mass of "Step 6", proceed to the following "Step 8".

Processing mark



8. If a few more samples are added, the processing indicator turns on. To prevent an error, add three or more. The processing indicator does not turn on if overloaded. Try to add the same number of samples as displayed.
9. The balance re-calculates the unit mass while the processing indicator is blinking. Do not touch the balance or samples on the pan until the processing indicator turns off.
10. Counting accuracy is improved when the processing indicator turns off.
11. Each time the above operation is performed, a more accurate unit mass will be obtained. There is no definite upper limit of ACAI range for the number of samples exceeding 100. Try to add the same number of samples as displayed.
12. Remove all the samples used in ACAI and proceed with the counting operation using the improved unit mass.


Note ACAI will not function on the unit mass entered using the keys, or digital input mode.

Storing the unit mass

By using the data memory function, 50 instances of storing a sample unit mass can be stored.

1. Set the function setting item "Data memory function (DATA)" to "Stores unit mass in counting (DATA)". Refer to "10.Function Table".
2. The displayed "P * *" is the selected unit mass registration number.
3. Press and hold the **PRINT** key to switch to the mode to change the unit mass registration number.

RE-ZERO key	Changes the registration number(+)
MODE key	Changes the registration number (-)
PRINT key	Decides on the displayed registration number.
CAL key	Cancel the displayed registration number.



4. Multiple unit masses can be stored by registering them with different unit mass registration numbers.

Note

- ※ Unit weight can be read by "U N:mm" command.
(mm corresponds to P01 to P50 with 01 to 50.)
- ※ The read unit mass can output by "?UW" command and can be changed by "?UW" command.

Note

- ※ ACAI cannot be used for the read unit mass.

5-3 Percent Mode (%)

The percent mode displays the weighting value in percentage compared to a 100% reference mass and is used for target weighing or checking the sample variance.

Selecting The Percent Mode

1. Press the **MODE** key to select the unit **%** (Percent mode).

Storing The 100% Reference Mass

2. Press the **SAMPLE** key to enter the 100% reference mass storing mode.
Even in the storing mode, pressing the **MODE** key will switch to the next mode.
3. Place a container on the weighing pan, if necessary. Press the **RE-ZERO** key to cancel the weight (tare). The balance displays **100 0 %**.
4. Place the sample to be set as the 100% reference mass on the pan or in the container.
5. Press the **PRINT** key to store the reference mass. The balance displays **100.00 %**. (The decimal point position depends on the reference value. The reference mass stored, even if the AC adapter is removed, is maintained in non-volatile memory.)

Note

- If the balance judges that the mass of the sample is too light to be used as a reference, it displays **Lo**.
- The displayed percentage is based on the 100% reference mass.


	100% mass	Decimal point position
Minimum display 0.001g model	0.100g ~ 0.999g	1%
	1.000g ~ 9.999g	0.1%
	10.000g ~	0.01%
Minimum display 0.01g model	1.00g ~ 9.99g	1%
	10.00g ~ 99.99g	0.1%
	100.00g ~	0.01%
Minimum display 0.1g model	1.0g ~ 9.9g	1%
	10.0g ~ 99.9g	0.1%
	100.0g ~	0.01%

- Registered values are stored even when the power is turned off.

6. Remove the sample

Reading the percentage

7. Please a sample to be compared to the reference mass on the pan.
The displayed percentage is based on the 100% reference mass.



5-4 Animal Weighing Mode (Hold Function)

This is the mode to weigh a moving object such as an animal, even when the display of the weighing data fluctuates. The hold function allows the average weight of the animal to be displayed. To use the hold function, set the function in the function table. Refer to "10. Function Table" and "10-3. Description Of The Class "Environment, Display" " for details.

6. Impact Detection Function





The GX-A / GF-A series has a function to detect impact to the mass sensor section and to display the impact level.

By lowering the impact level at the time of loading, it is possible not only to alleviate variation in the weighing value but also to reduce the risk of failure of the mass sensor section.

Especially when incorporating the balance in a production line, etc. and weighing by means such as an automated system, impact to the sensor may be applied greater than expected.

When designing automatic systems and the like, it is recommended that you minimize the impact level as much as possible while checking the shock indicator.

Impact level display is from level 0 to level 4, 5level.

Impact level	Shock indicator	Buzzer	Contents
0	No	No	Safe
1	SHOCK 	No	Caution
2	SHOCK 	No	Caution : Consider impact mitigation
3	SHOCK 	One beep	Warning : Do not apply greater impact
4	SHOCK 	Two beep	Danger : Sensor may be damaged



Note

- Impact on the weighing sensor may be applied to the weighing pan at time of loading, or it may be applied from the table on which the balance is installed.

The impact detection function also works for impact applied from the table.

7. Response Adjustment / Self Check Function

This function stabilizes the weight value, reducing the influence on weighing that is caused by drafts and/or vibration at the place where the balance is installed. This function adjusts by automatically analyzing the environment or by hand-operation. The function has three stages as follows : Changing the weighing speed changes the display refresh rate.


Display	Function setting	Response characteristic	
FAST	[Cond 0]	Fast response, 	Sensitive value 
MID.	[Cond 1]		
SLOW	[Cond 2]	Slow response,	Stable value



7-1 Response Adjustment

Response adjustment can be changed by the following method.

1. Press and hold the [MODE] key until [RESPONSE] is displayed, and then release the key.
2. Press the [MODE] key to select a weighing speed. Either [FAST], [MID] or [SLOW] can be selected.
3. After a few seconds of inactivity the balance displays [End]. Then, it returns to the weighing mode and displays the updated response indicator.
The response indicator remains displayed for a while.



Note

The value set for response adjustment can be changed by "Condition [Cond]" of "Environment display" in the Function Table.

For the setting method, refer to "10.Function Table".

7-2 Self-Check-Function / Automatic Setting Of Minimum Weight Value

With the self-diagnosis function confirmation and display of repeatability can be performed in addition to failure diagnosis, and whether or not the balance's performance is being exhibited can be easily checked.

It is also possible to display and register the reference value of the minimum weighing value using repeatability data.

For details of the minimum weighing value, refer to the technical information on our website.

(http://www.aandd.co.jp/adhome/products/balance/tecdoc_balance.html)

1. Press and hold the **MODE** key while weighing is displayed.
2. Release the key when **RESPONSE** display blinks.
3. **CH** displays and the self-diagnosis function is started.
4. When the diagnosis is completed, the diagnosis result is displayed.

When there are no problems in the balance, the **CH PASS** display blinks.

If **CH FAIL** is displayed blinking, there is a possibility that a fatal fault has occurred in the balance. Please request repair.

SAMPLE key It is possible to switch the display of diagnostic result, repeatability, minimum weight value.

PRINT key The displayed contents are output.

5. Press the **CAL** key to display **End** and move to the display for minimum weighing value registration.
6. When registering as minimum weighing value of minimum weight value display warning function performs the following.

Press the **RE-ZERO** key to change No/Go and display **Go**.


Press the **PRINT** key.

If not registered, press the **PRINT** key while displaying **No**, No change or press **CAL**. Return to the weighing display.

* For minimum weighing warning function, refer to " 16.Minimum Weighing Warning Function".

Note

For the minimum weighing warning function, settings can only be made when logged in as administrator (AdMin) if the password lock function is turned on.



8. Calibration

Since the balance's resolution is high, weighing values may change due to gravity and daily environmental changes. It is necessary to perform calibration (sensitivity adjustment) with the weight in order to keep the weighing values from changing even if gravity or the environment changes.

It is recommended that you calibrate if the balance is installed for the first time or relocated, or when the weighing values change significantly in daily inspection, etc.

Adjustment means to adjust the weighing value of the balance using the reference weight or internal mass. Calibration is to weigh with the reference weight and compare how much the result deviates from the reference value. (Adjustment is not performed in calibration.)

Calibration (Sensitivity adjustment)

- Auto calibration-----Automatically adjust the balance using the internal mass depending on the temperature change of the operating environment or the set time and interval time. (GX-A series)
- Calibration test using the internal mass -----Using the internal mass, adjust the balance with a single touch.
- Calibration test using an external weight-----Using an external mass, adjust the balance with an external mass.

Calibration test (Sensitivity calibration)

Calibration test with your own weight-----Output the result of checking the accuracy of weighing using your own weight.

*No adjustment is made.

Caution

- Do not allow vibration or drafts to affect the balance during calibration.
- To output the data for GLP/GMP using the RS-232C interface, set "GLP/GMP output (INF0)" of "Data output (dout)". Refer to "10. Function Table". The time and date can be added to the GLP/GMP report. If the time or date is not correct, adjust them. Refer to "10-4 Clock and Calendar Function".
- The calibration and calibration test data can be stored in memory. To store them, set "Data memory (dMEm)" .
- The accuracy of the weight used in calibration affects the accuracy of the balance after calibration.
- Select the mass to be used for calibration and calibration tests from the table below.

Model	Usable calibration weight	Factory setting	Adjustable range
GX-203A, GF-203A	50g, 100g, 200g	200g	-9.999g ~ +9.999g
GX-303A, GF-303A	50g, 100g ~ 300g (100g interval)	200g	
GX-403A, GF-403A	50g, 100g ~ 400g (100g interval)	400g	
GX-603A, GF-603A	50g, 100g ~ 600g (100g interval)	500g	
GX-1003A, GF-1003A	50g, 100g ~ 1000g (100g interval)	1000g	
GX-1603A, GF-1603A	50g, 100g ~ 1600g (100g interval)	1000g	
GX-2002A, GF-2002A	500g, 1000g, 2000g	2000g	-99.99g ~ +99.99g
GX-3002A, GF-3002A	500g, 1000g ~ 3000g (1000g interval)	2000g	
GX-4002A, GF-4002A	500g, 1000g ~ 4000g (1000g interval)	4000g	
GX-6002A, GF-6002A	500g, 1000g ~ 6000g (1000g interval)	5000g	
GX-10002A, GF-10002A	500g, 1000g ~ 10000g (1000g interval)	10000g	
GX-6001A, GF-6001A	500g, 1000g ~ 6000g (1000g interval)	5000g	-999.9g ~ +999.9g
GX-10001A, GF-10001A	500g, 1000g ~ 10000g (1000g interval)	10000g	

Display



This indicator means "In process of measuring calibration data".

Do not allow vibration or drafts to affect the balance while the indicator is displayed.

8-1 Automatic Self Calibration For The GX-A Series

This function automatically calibrates the balance when the balance detects an ambient temperature change. If GLP output is selected in the function table, the balance outputs the calibration report or stores the data in memory.

In the auto calibration mode, either the temperature change, the setting time, or the interval time can be set with the function setting $[Fnc]$.

For the setting time, the three function setting of $[TIME1]$, $[TIME2]$ and $[TIME3]$ can be set.

Interval time can be set from 0.5h to 24h with function setting $[int]$.

Caution

If something is on the weighing pan, the balance judges that it is in use and does not perform automatic self calibration.


The criteria that the balance judges is in use are as follows.

0.001g models	0.01g models	0.1g models
Lower than 2g	Lower than 20g	Lower than 20g

To maintain the calibrated state, keep the weighing pan clear while not in use.



Indicates that the balance detects a change in ambient temperature and automatic self calibration will start. If the balance is not used for a few minutes with this indicator (◀) blinking, the balance performs automatic self calibration. The blinking duration depends on the environment.



Indicates that the balance is measuring calibration data. Do not allow vibration or drafts to affect the balance while this indicator is displayed. After calibration, the balance returns to indicate the previous display.

Note The balance can be used while the indicator blinks. But, it is recommended that to maintain the accuracy, stop using the balance and confirm that there is nothing on the pan and allow the balance to perform self calibration.

8-2 One-Touch Calibration For The GX-A Series

This function calibrates the balance using the internal mass.

- Step1 Connect the AC adapter and warm up the balance for at least 30 minutes with nothing on the weighing pan.
- Step2 Press the **CAL** key. The balance displays **CRL in** and performs calibration using the internal mass. Do not allow vibration or drafts to affect the balance.
- Step3 The balance displays **End** after calibration. If the "GLP output (inFd)" parameter of the function table is set to "1" or "2", the balance displays **GLP** and outputs the "calibration report" using the RS-232C interface or stores the data in memory. Refer to "11-3. GLP Report" and "Data memory (dRR)" of the function table for details.
- Step4 The balance will automatically return to the weighing mode after calibration.

About the internal mass


The value of the internal mass may change due to aging, corrosion or other damage caused by the operating environment. Check the internal mass periodically. Correct the internal mass value as necessary. Refer to "8-5. Correcting the internal mass value".

To maintain the weighing accuracy, perform the calibration using an external weight periodically, as described below.

8-3 Calibration Using An External Weight

This function calibrates the balance using an external weight.

- Step 1 Connect the AC adapter and warm up the balance for at least 30 minutes with nothing on the weighing pan.
- Step 2 Press and hold the **CAL** key until **CAL out** is displayed, then release the key.
- Step 3 If you want to change the setting of the external weight value, press the **SAMPLE** key and go to step 4.
If you don't change the setting of the external weight value, go to step 5.
- Step 4 Specify the calibration weight value as follows:
SAMPLE key To switch the display condition to: "All of the segments blinking" (calibration weight selection mode) or "The last four digits blinking" (value adjustment mode).
RE-ZERO key To select the calibration weight or adjust the value.
MODE key In the value adjustment mode, -9999 digits appear after +9999 digits.
PRINT key To store the new weight value. Even if the AC adapter is removed, the data is maintained in non-volatile memory.
CAL key To cancel the operation and return to **CAL 0**.
- Step 5 Confirm that there is nothing on the pan and press the **PRINT** key. The balance measures the zero point. Do not allow vibration or drafts to affect the balance. The balance displays the calibration weight value.
- Step 6 Place the displayed calibration weight on the pan and press the **PRINT** key. The balance measures the calibration weight. Do not allow vibration or drafts to affect the balance.
- Step 7 Remove the weight from the pan.
- Step 8 After calibration, if GLP output is to be set, "Calibration report" is output or stored in data memory.
- Step 9 The balance will automatically return to the weighing mode.
- Step 10 Place the calibration weight on the pan and confirm that the value displayed is within ± 2 digits of the specified value. If it is not within the range, check the ambient conditions such as breeze and vibration also check the weighing pan. Then, repeat steps 1 to 10.



8-4 Calibration Test Using An External Weight

This function tests the weighing accuracy using an external weight and outputs the result. This is available only when the GLP output parameter is set to " *dout info 1* ". (Calibration test does not perform calibration)

Step1 Connect the AC adapter and warm up the balance for at least 30 minutes with nothing on the weighing pan.

Step2 Press and hold the **CAL** key until **CCout** is displayed

Step3 Release the key.

Step4 If you want to change the setting of the external weight value, press the **SAMPLE** key and go to step 5.
If you don't change the setting of the external weight value, go to step 6.

Step5 Specify the calibration weight value as follows:

SAMPLE key To switch the display condition to: "All of the segments blinking" (calibration weight selection mode) or "The last four digits blinking" (value adjustment mode).

RE-ZERO key To select the calibration weight or adjust the value. In the value adjustment mode, -9999 digits appear after +9999 digits.

PRINT key To store the new weight value. Even if the AC adapter is removed, the data is maintained in non-volatile memory.

CAL keyTo cancel the operation and return to **CC 0**.

Step6 Confirm that there is nothing on the pan and press the **PRINT** key. The balance measures the zero point.


Do not allow vibration or drafts to affect the balance.

The balance displays the calibration weight value.

Step7 The balance displays the zero point measured value several seconds.
Replace the external weight.

Step8 The balance outputs or stores "calibration test data".

Step9 The balance will automatically return to the weighing mode.



8-5 Correcting The Internal Mass Value Of The GX-A series

Internal mass value can be corrected with function setting [5] 17.

There are two correction methods, as follows.

Auto····· This is a method of correcting the internal mass weight value based on an external weight.

Manual···· This is a method of correcting by digitally inputting a correction reference value (internal weight conversion value).

Note

- Correction of internal mass value can not be executed at factory setting.


Refer to "9. Function Switch And Initialization" or the following setting method, and enable changing of the function setting and correction the internal mass value.

Setting procedure


1. Press the **ON:OFF** key to turn off the display.
2. Hold down the **PRINT** and **SAMPLE** keys, and press the **ON:OFF** key to display **P5**.
3. Press the **PRINT** key and set the "internal mass correction switch" and "function setting switch" to "1" with the next key.

SAMPLE key Select the switch (blinking digit).

RE-ZERO key Change the value of the blinking switch.



4. Press the **PRINT** key to register and display the weighing display.



8-6 Correcting The Internal Mass Value Of The GX-A series (Auto)

Calibrate referring to "8-3. Calibration Using An External weight".


This is method of correcting the internal mass weight value based on an external weight. After calibration with the external mass, the balance automatically loads and unloads the internal mass and corrects the internal mass value.

The available masses are as shown in the table below. The corrected mass value is maintained in non-volatile memory even if the AC adapter is removed.

Model	Available mass	Factory setting	Adjustable range
GX-203A, GF-203A	50g, 100g, 200g	200g	-9.999g ~ +9.999g
GX-303A, GF-303A	50g, 100g ~ 300g (100g interval)	200g	
GX-403A, GF-403A	50g, 100g ~ 400g (100g interval)	400g	
GX-603A, GF-603A	50g, 100g ~ 600g (100g interval)	500g	
GX-1003A, GF-1003A	50g, 100g ~ 1000g (100g interval)	1000g	
GX-1603A, GF-1603A	50g, 100g ~ 1600g (100g interval)	1000g	
GX-2002A, GF-2002A	500g, 1000g, 2000g	2000g	-99.99g ~ +99.99g
GX-3002A, GF-3002A	500g, 1000g ~ 3000g (1000g interval)	2000g	
GX-4002A, GF-4002A	500g, 1000g ~ 4000g (1000g interval)	4000g	
GX-6002A, GF-6002A	500g, 1000g ~ 6000g (1000g interval)	5000g	
GX-10002A, GF-10002A	500g, 1000g ~ 10000g (1000g interval)	10000g	
GX-6001A, GF-6001A	500g, 1000g ~ 6000g (1000g interval)	5000g	-999.9g ~ +999.9g
GX-10001A, GF-10001A	500g, 1000g ~ 10000g (1000g interval)	10000g	

Setting procedure


- The internal mass value cannot be corrected at factory settings. Refer to "9. Function Switch and Initialization" and enable changing of the function setting and correction the internal mass value.
- In weighing mode, press and hold the **SAMPLE** key to display **bASFnC**.
- Press the **SAMPLE** key several times until **CS in** appears.
- If **CS in** does not display, perform Step 1.
- Press the **PRINT** key to display **Auto**.
- When preparation is completed, press the **PRINT** key.
- CAL SET** is displayed and the internal mass value is automatically corrected.
- When adjustment of the internal mass value is completed, **CAL in** is displayed and calibration is performed automatically with the adjusted internal weight.
- End** is displayed and you are returned to the weighing mode.
- Place the external weight used for calibration on the balance check whether the balance was corrected. If it is not corrected properly, return to Step 2.



8-7 Correcting The Internal Mass Value Of The GX-A series (Manual)

The balance can correct the internal mass value within the range shown below. This function corrects the internal mass value to conform to an external weight. The corrected mass value is maintained in non-volatile memory even if the AC adapter is removed. The internal mass value is corrected as follows:

Model	Target	Range	Model	Target	Range
GX-203A	200.000g	±9.999g	GX-2002A	2000.00g	±99.99g
GX-303A	200.000g	±9.999g	GX-3002A	2000.00g	±99.99g
GX-403A	200.000g	±9.999g	GX-4002A	2000.00g	±99.99g
GX-603A	500.000g	±9.999g	GX-6002A	5000.00g	±99.99g
GX-1003A	1000.000g	±9.999g	GX-10002A	5000.00g	±99.99g
GX-1603A	1000.000g	±9.999g	GX-6001A	5000.0g	±999.9g
			GX-10001A	5000.0g	±999.9g




After performing one touch calibration, place the external weight on the balance and check the correction amount.

(In the example, since there is deviation of -0.06g at 2000g, the correction amount for the GX-6002A is +0.015g/5kg because the correction target is 5000g)

Setting procedure

- Press and hold the **[SAMPLE]** key to display **bASFnC**.
(Enter the function setting)
- Press the **[SAMPLE]** key several times until **CS 17** appears.
- Press the **[PRINT]** key to display **Auto**.
- Press the **[SAMPLE]** key to display **MANUAL** and press the **[PRINT]**.
Select the following keys.
 - [RE-ZERO]** key(+) Select the correction value.
(After +9999 digits will be -9999 digits.)
 - [MODE]** key(-) Select the correction value.
(After -9999 digits will be +9999 digits.)
 - [PRINT]** key Register and display the following items.
 - [CAL]** key Cancel and display the following items.
- Press the **[CAL]** key twice, to return to the weighing display.
- Press the **[CAL]** key and calibrate with the internal mass.
- Place the external weight on the balance check whether the value was corrected properly. If it is not corrected properly, return to Step 1.



9. Function Switch And Initialization

9-1 Permit Or Inhibit


The balance stores parameters that must not be changed unintentionally adjustment data for accurately weighing, data for adapting to the usage environment, data to control the communications interface, etc. "A function selection switch" is provided to protect those parameters and it can be used to select "change prohibited" or "changeable (usable)". By setting to "change prohibited", that function cannot be entered, so inadvertent change.

"Switch for function selection" has the following five.

"Function table", "Calibration using the internal mass", "Calibration using the external weight", "Automatic self calibration", "Internal mass correction".

- Step1 Press the **ON:OFF** key to turn off the display.
- Step2 While pressing and holding the **PRINT** key and the **SAMPLE** key, press the **ON:OFF** key to display **PS**.
- Step3 Press the **PRINT** key. Then the balance displays the function switches.
- SAMPLE** key To select a switch to change the parameter. The selected switch blinks.
 - RE-ZERO** key To change the parameter of the switch selected.
 - To inhibit changes. (Can not be used.)
 - To permit changes. (Can be used.)
 - PRINT** key To store the new parameter and return to the weighing mode.
 - CAL** key To cancel the operation (display **Err**) . Press the **CAL** key and return to the weighing mode


Example of GX-A series



Example of GF-A series

R-00101

The display shown left indicates the factory settings.




9-2 Initializing The Balance

This function returns the following parameters to factory settings. Calibration data

- Function table
- The sample unit mass value (counting mode),
100% reference mass value (percent mode)
- The data that is stored in the balance using the data memory function
- External calibration weight and target weight value
- Function switch settings

Note Be sure to calibrate the balance after initialization.

- Step1 Press the **ON:OFF** key to turn off the display.
- Step2 While pressing and holding the **PRINT** key and the **SAMPLE** key, press the **ON:OFF** key to display **P5**.
- Step3 Press the **SAMPLE** key to display **[Lr]**.
- Step4 Press the **PRINT** key.
To cancel this operation, press the **CAL** key.
- Step5 Press the **RE-ZERO** key to change **N0 / 50**.
- Step6 With displaying **[Lr 50]**, press the **PRINT** key to initialize the balance. The balance will automatically return to the weighing mode.









10. Function Table

The function table reads or rewrites the parameters that are stored in the balance. These parameters are maintained in non-volatile memory, even if the AC adapter is removed.

The function table menu consists of two layers. The first layer is the "Class" and the second layer is the "Item".

10-1 Setting The Function Table

Display symbol and keys


	The symbol "●" shows effective parameter.
	When pressing and holding the key in the weighing mode, the balance enters the function table mode. The key to select the class or item in the function table mode.
	The key to change the parameter.
	The key to change the parameter.
	When a class is displayed, moves to an item in the class. When an item is displayed, stores the new parameter and displays the next class.
	When an item is displayed, cancels the new parameter and displays the next class. When a class is displayed, exits the function table mode and returns to the weighing mode.

Setting procedure

- Step 1 Press and hold the **SAMPLE** key until **bRSFnC** of the function table is displayed in the weighing mode, then release the key
- Step 2 Press the **SAMPLE** key to select a class.
- Step 3 Press the **PRINT** key to enter the class
- Step 4 Press the **SAMPLE** key to select a item.
- Step 5 Press the **RE-ZERO** key to select a parameter for the selected item.
- Step 6 To change another (multiple) item with the same class, repeat Step 4 and Step 5. To end the setting change of the same class, proceed to Step 7.
- Step 7 If storing parameters of the selected class, press the **PRINT** key.
Then the next class is displayed.
If canceling the current operation, press the **CAL** key. Then the next class is displayed.
- Step 8 When specifying parameters for another class, proceed to Step 2.
When finishing the setting, press the **CAL** key to return to weighing mode.

Setting Example

This example sets "Stores weighing data" for "Data memory" and "1 minute" for "Interval time".



10-2 Details Of The Function Table

Class	Item	Parameter	Description		
bR5Fnc Environment Display	Cond Condition	0	Fast response, sensitive value	Can be changed by response adjustment. With "Hold 1", sets the averaging time.	
		▪ 1	↕		
		2			Slow response, stable value
	St-b Stability band width	0	Stable when within ± 1 digit	↕	The stabilization indicator illuminates with the display fluctuation within the range. With "Hold 1", sets the stabilization range.
		▪ 1	Stable when within ± 3 digit		
		2			
	Hold Hold function	▪ 0	OFF	Holds the display when stable in animal mode. With "Hold 1", [ANIMAL] turns on.	
		1	ON		
	trc Zero tracking	0	OFF	Keeps zero display by tracking zero drift.	
		▪ 1	Normal		
		2	Strong		
		3	Very strong		
	SPd Display refresh rate	▪ 0	5 times / second	Output frequency approx. 5.21Hz	
		1	10 times / second	Output frequency approx. 10.42Hz	
		2	20 times / second	Output frequency approx. 20.83Hz	
Pnt Decimal point	▪ 0	Point (.)	Decimal point format		
	1	Comma (,)			
P-on Auto display-ON	▪ 0	OFF	Turns on the weighing mode display when AC adapter is connected		
	1	ON			
P-off Auto display-OFF	▪ 0	OFF	Turns off the display after 10 minutes of inactivity.		
	1	ON			
rn0 Minimum display	▪ 0	Display minimum display digit	Display at weighing start		
	1	Not display minimum display digit			
bEEP Buzzer	0	OFF			
	▪ 1	ON			
P-ZERO Stores tare value	▪ 0	OFF Zero indication at power on	ON Previous time weighing indication at power on		
	1	ON			
dISP-LED Backlight brightness	0~9	10%~100%			
	▪ 5	Factory setting			
LV-LED Bubble spirit level lightning	0	OFF	Bubble spirit level LED lightning		
	▪ 1	ON			
CL Add Clock		See "10-4 Clock and calendar Function"		Confirms and sets the time and date. The time and date are added to output data.	

▪ Factory setting

Note: "Digit" is a unit of minimum weighing value.

Class	Item	Parameter	Description		
[P] Fnc Comparator	[P] Comparator mode	▪ 0	No comparison		
		1	Comparison when stable value or overloaded		
		2	Continuous comparison		
	[P-t] Number of comarator stages	▪ 0	3 stage comparator	HI, OK, LO	
		1	5 stage comparator	HH, HI, OK, LO, LL	
	[P-z] Near zero	0	Also comare near zero		
		1	± 5 are not compared		
		▪ 2	± 10 are not compared		
		3	± 20 are not compared		
		4	± 50 are not compared		
		5	± 100 are not compared		
	[P-p] Polarity	0	Plus only		
		1	Minus only		
		▪ 2	Bipolarity		
	[P in] Input method	▪ 0	Digital input, upper / lower limits		
1		Weighing input, upper / lower limits			
2		Digital input, reference value			
3		Weighing input, reference value			
[P Frd] Flow measurement	▪ 0	Comparison by flow rate value			
	1	Comparison by weighing value			
[P] VALUE Comparator value	[P HH] Second upper limit	See "10-5 comparator Function"		Displayed only when [P in] is set to digital input. [P HH], [P LL] are displayed only when 5step comparator is set.	
	[P HI] Upper limit				
	[P LO] Lower limit				
	[P LL] Second lower limit				
	[P REF] Reference value	See "10-5 comparator Function"		Displayed only when [P in] is set by input by load. [P LME2] is displayed only when 5 step comrator is set.	
	[P LME] Tolerance value				
	[P LME2] Second tolerance value				

▪ Factory setting

Note: "Digit" is a unit of minimum weighing value.

Class	Item	Parameter	Description		
CP bEEP Comparator buzzer	bEP HH HH buzzer	<ul style="list-style-type: none"> ▪ 0 OFF ▪ 1 ON 		Displayed only when 5 step comparator is set.	
	bEP HI HI buzzer	<ul style="list-style-type: none"> ▪ 0 OFF ▪ 1 ON 			
	bEP OK OK buzzer	<ul style="list-style-type: none"> ▪ 0 OFF ▪ 1 ON 			
	bEP LO LO buzzer	<ul style="list-style-type: none"> ▪ 0 OFF ▪ 1 ON 			
	bEP LL LL buzzer	<ul style="list-style-type: none"> ▪ 0 OFF ▪ 1 ON 		Displayed only when 5 step comparator is set.	
dout Data output	Prt Data output mode	<ul style="list-style-type: none"> ▪ 0 Key mode 		Accepts the PRINT key only when the display is stable.	
		<ul style="list-style-type: none"> ▪ 1 Auto print mode A (Reference=zero) 		Outputs data when the weighing value stabilizes beyond the range from $RP-P$ to $RP-b$ from the zero point.	
		<ul style="list-style-type: none"> ▪ 2 Auto print mode B (Reference=last stable value) 		Outputs data when the weighing value stabilizes beyond the range from $RP-P$ to $RP-b$ from last stable value.	
		<ul style="list-style-type: none"> ▪ 3 Stream mode 		Outputs data at the specified display refresh rate.	
		<ul style="list-style-type: none"> ▪ 4 Key mode B (Immediately) 		Accepts the PRINT key regardless of the display condition.	
		<ul style="list-style-type: none"> ▪ 5 Key mode C (When stable) 		Accepts the PRINT key immediately when the display is stable, or waits for the display to be stable when not.	
		<ul style="list-style-type: none"> ▪ 6 Interval output mode 		Outputs data for each time set by int .	
	RP-P Auto print polarity	<ul style="list-style-type: none"> ▪ 0 Plus only ▪ 1 Minus only ▪ 2 Biolarity 		Regardless of displayed value	
		RP-b Auto print difference	<ul style="list-style-type: none"> ▪ 0 10 digit ▪ 1 100 digit ▪ 2 1000 digit 		Difference between reference value and displayed value

▪ Factory setting

Note: "Digit" is a unit of minimum weighing value.

Class	Item	Parameter	Description	
<i>dout</i> Data output mode	<i>DATA</i> Data memory	0	OFF	Refer to "12*Data Memory"
		1	Stores unit mass in counting mode	
		2	Stores the weighing data and calibration history	
	<i>int</i> Interval time	0	Every measurement	Interval time in the interval memory mode when using <i>PrtE</i> .
		1	2 seconds	
		2	5 seconds	
		3	10 seconds	
		4	30 seconds	
		5	1 minutes	
		6	2 minutes	
		7	5 minutes	
		8	10 minutes	
	<i>d-no</i> Data number	0	No output	Valid when data memory function is ON.
		1	Output	
	<i>S-tD</i> Time/Date output	0	No output	Refer to "10-4 Clock And Calender Function"
		1	Time only	
		2	Date only	
		3	Time and date	
	<i>S-ID</i> ID number output	0	No output ID number	
		1	Output ID number	
<i>PUSE</i> Data output pause	0	OFF	Selects the data output interval.	
	1	ON open 1.6 seconds		
<i>At-F</i> Auto feed	0	OFF	Selects whether or not automatic feed is performed.	
	1	ON open 1 line		
<i>inFo</i> GLP output	0	OFF	Refer to "10-3 GLP output"	
	1	ON		
<i>Rr-d</i> Zero after output	0	OFF	Function to apply re-zero after outputting data.	
	1	ON		
<i>UFC</i> UFC function	0	OFF	Refer to "Communication manual" on the A&D website.	
	1	ON		

■ Factory setting

Note: "Digit" is a unit of minimum weighing value.

Class	Item	Parameter	Description		
Serial interface	Mode Access point	0	PC	All communication setting are possible.	
		1	Printer	Only TYPE 0,1 can be selected.	
		2	External indicator	Selects stream with TYPE 0.	
	bps Baud rate	0	600bps		
		1	1200bps		
		2	2400bps		
		3	4800bps		
		4	9600bps		
		5	19200bps		
	bPr Data bit, parity bit	0	7 bit EVEN		
		1	7 bit ODD		
		2	8 bit None		
	CrLF Terminator	0	CR LF		CR: ASCII 0Dh code
		1	CR		LF : ASCII 0Ah code
	TYPE Data format	0	A&D standard format		Refer to "communication manual " on the A&D website.
1		DP format			
2		KF format			
3		MT format			
4		NU format			
t-UP Command time out	0	No limited		Selects wait time during command reception	
	1	Limited for one second			
ErEd AK, error code	0	OFF		AK: ASCII 06h code	
	1	ON			
USB interface	UFnc USB action mode	0	Quick USB ALL	Refer to "communication manual " on the A&D website.	
		1	Quick USB NU		
		2	Bidirectional USB virtual COM	insallation of dedicated driver is required for PC.	
	U-EP USB data format	0	A&D standard format	Refer to "communication manual " on the A&D website.	
		1	NU format		
		2	CSV format		
		3	TAB format		

▪ Factory setting

Note: "Digit" is a unit of minimum weighing value.

Class	Item	Parameter	Description		
<i>AP Fnc</i> Application function	<i>APP</i> Application mode	0	Normal weighing mode		
		▪ 1	Capacity indicator		
		2	Statistical calculation mode		
		3	Flow measurement mode		
		4	Gross, Net, Tare mode		
	<i>STATF</i> Statistical function mode output items.	▪ 0	Number of data, sum		
		1	Number of data, sum, max, min, range(max-min), average		
		2	Number of data, sum, max, min, range(max-min), average, standard deviation, coefficient of variation		
		3	Number of data, sum, max, min, range(max-min), average, standard deviation, coefficient of variation, relative error		
	<i>FRD Unit</i> Flow rate unit	▪ 0	g / s (gram/second)		Refer to "14. Flow Measurement"
		1	g / m (gram/minute)		
		2	g / h (gram/hour)		
		3	mL / s (milliliter/second)		
		4	mL / m (milliliter/minute)		
		5	mL / h (milliliter/hour)		
<i>Calc Auto</i> Calculation time automatic setting	▪ 0	OFF			
	1	ON			
<i>MW Fnc</i> Minimum Weighing Warning Function	<i>MW-CP</i> Minimum weighing comparison	0	No comparison Do not use MW Fnc		
		▪ 1	Comparison without near zero		
	<i>MW</i> Minimum weighing value input	2	Comparison including near zero		
		Refer to "15. Minimum Weighing Warning Function"			
<i>Unit</i> Unit		Refer to "4. Weighing Units"			
<i>dS Fnc</i> Specific gravity measuring function	<i>Ld in</i> Liquid density input	▪ 0	Water temperature		
		1	Liquid density		
	<i>dS</i> Specific gravity measuring mode	▪ 0	Density measurement of a soil		
		1	Density measurement of a liquid		
<i>id</i> ID number setting		Refer to "11-2 Setting The ID Number"			

▪ Factory setting

Note: "Digit" is a unit of minimum weighing value.

Class	Item	Parameter	Description		
<i>PASSwd</i> * Password lock	<i>LOCK</i> Lock function	▪ 0	OFF	Refer to "19. Password Lock Function"	
		1	ON		
	<i>PASSNo.</i> Password registration	<i>Admin.</i>	Administrator password input		
		<i>OPR. 1</i> }	Operator 1 password input		
<i>OPR. 10</i>		Operator 10 password input			
<i>AutoCAL</i> Auto calibration	<i>FUNC</i> Calibration mode	▪ 0	Setting temperature		
		1	Setting time		
		2	Interval time		
	<i>TIME 1</i> Setting time1	Refer to "8-1 Automatic Self calibration"			
	<i>TIME 2</i> Setting time2				
	<i>TIME 3</i> Setting time3				
	<i>Int</i> Interval time				
	<i>CS in</i> * Correction of internal mass value.	<i>Auto</i>	Automatic input	Refer to "8-5 Correcting The Internal Mass Value Of The GX-A series Auto	
<i>MANUAL</i>		Digital input of correction value	Refer to "8-6 Correcting The Internal Mass Value Of The GX-A series Manual		

* is GX-A series only.

▪ Factory setting

Note: "Digit" is a unit of minimum weighing value.

10-3 Description Of The Class "Environment, Display"

Condition ($Cond$)

- $Cond \ 0$ This parameter is for sensitive response to the fluctuation of a mass value. Used for powder target mass, weighing a very light sample or when quick response weighing is required. After setting, the balance displays **FAST**.
- $Cond \ 2$ This parameter is for stable weighing with slow response. Used to prevent a mass value from drifting due to vibration or drafts. After setting, the balance displays **SLOW**.

Stability band width ($St-b$)

This item controls the width to regard a mass value as a stable value. When the fluctuation per second is less than the parameter, the balance displays the stabilization indicator and outputs or stores the data by function setting (*count, dATA* etc.) The parameter influences the "Auto print mode". Also, the minimum display being displayed is 1 digit.

ex. If 0.01 mg display is selected by pressing the **SAMPLE** key on the GX-303A, 0.01 mg is 1 digit.

- $St-b \ 0$ This parameter is used for sensitive response of the stabilization indicator. Used for exact weighing.
- $St-b \ 2$ This parameter ignores slight fluctuations of a mass value. Used to prevent a mass value from drifting due to vibration or drafts

Hold function ($Hold$) (Animal weighing mode)

This function is used to weigh a moving object such as an animal. When the weighing data is over the weighing range from zero and the display fluctuation is within the stabilization range for a fixed period of averaging time, the processing indicator illuminates and the balance displays the average weight of the animal. When the animal is removed from the weighing pan, the display returns to zero automatically. This function is available only when the hold function parameter is set to "I" (the animal mode indicator **ANIMAL** illuminates) and any weighing unit other than the counting mode is selected. The stabilization range and averaging time are set in "Condition ($Cond$)" and "Stability band width ($St-b$)".

Weighing range		Averaging time		Stabilisation range	
0.001g model	0.200g or more	$Cond \ 0$	2sec.(Efficiency priority)	$St-b \ 0$	Lesser 6.25%
0.01g model	2.00g or more	$Cond \ 1$	4sec.	$St-b \ 1$	12.5%
0.1g model	20.0g or more	$Cond \ 2$	8sec.(Exact priority)	$St-b \ 2$	Greater 16.7%

※ Animal container kit (GXA-12) can be installed except GX-203A.

Zero tracking (trc)

This function tracks zero point drift caused by changes in the environment and stabilizes the zero point. When the weighing data is only a few digits, turn the function off for accurate weighing.

- $trc \ 0$ The tracking function is not used. Used for weighing a very light sample.
- $trc \ 1$ The normal tracking function is used. (± 1 digit / 1 second)
- $trc \ 2$ The strong tracking function is used. (± 1 digit / 0.5 second)
- $trc \ 3$ The very strong tracking function is used. (± 2 digit / 0.2 second)

”

Display refresh rate (*SPd*)

The periodic time to refresh the display. This parameter influences "Baud rate", "Data output pause" and the data output rate of "Stream mode".

Decimal point (*Pnt*)

The decimal point format can be selected.

Auto display-ON (*P-on*)

When the AC adapter is plugged in, the display is automatically turned on without the ON:OFF key operation, to display the weighing mode. Used when the balance is built into an automated system. One hour warm up is necessary for accurate weighing.

Auto power-OFF (*P-off*)

This is a function to turn off only the display automatically when there is no operation made for a certain amount of time (approximately 10 minutes) while the power is on.

Minimum display (*rnL*)

When weighing with rough precision, the minimum display can be turned off without key operation. This is useful when built into an automated system/

Buzzer (*bEEP*)

Select ON/OFF for the built-in buzzer that sounds when a key is operated or the status changes.

Tare value record (*P-ZErD*)

After turning on the power supply, the display will not be automatically set to zero, and it will start from the previous weighing value. This is useful when a hopper, etc. is attached to the weighing pan and the power needs to be turned off while weighing discharge.

Backlight brightness (*d,SP-Ltd*)

Select the brightness of the backlight of the LCD display.

Bubble sprit level lightning (*LW-LED*)

Select ON/OFF for the LED that illuminates the bubble sprit level.

10-4 Clock And Calendar Function

The balance is equipped with a clock and calendar function.

When the Clock and Calendar function (dout, 5-td) is set, the time and date are added to the output data.

Set or confirm the time and date as follows:

Operation

Step1 Press and hold the **[SAMPLE]** key until **bASFnC** of the function table is displayed in the weighing mode, then release the key.

Step2 Press the **[SAMPLE]** key several times to display **[CL Adj]**.

Step3 Press the **[PRINT]** key.

The balance enters the mode to confirm or set the time and date.

Confirming the time

Step4 The current time is displayed with all the digits blinking.

- When the time is not correct and is to be changed, press the **[RE-ZERO]** key and go to step 5.
- When the time is correct and the date is to be confirmed, press the **[SAMPLE]** key and go to step 6.
- When the time is correct and the date does not need to be confirmed, press the **[CAL]** key and go to step 8.

Setting the time

Step5 Set the time in 24-hour format using the following keys.

[RE-ZERO](+)key..... To increase the value by one.

[MODE](-)key..... To decrease the value by one.

[SAMPLE] key..... To select the digits to change the value.
The selected digits blink.


[PRINT] key..... To store the new setting, display **[End]** and go to step 6.

[CAL] key..... To cancel the new setting and go to step 6

Confirming the date

Step6 The current date is displayed with all the digits blinking.

- To change the display order of year (Y), month (M) and day (D), press the **[MODE]** key. The date is output in the order as specified.
- When the date is not correct and is to be changed, press the **[RE-ZERO]** key and go to step 7.
- When the date is correct and the operation is to be finished, press the **[CAL]** key and go to step 8.
- When the time is to be confirmed again, press the **[SAMPLE]** key and go back to step 4.



Setting the date

Step7 Set the date using the following keys.

RE-ZERO(+)**key** To increase the value by one.

MODE(-)**key** To decrease the value by one.

SAMPLE **key** To select the digits to change the value.

The selected digits blink.

PRINT **key** To store the new setting, display **End** and go to step 8.

CAL **key** To cancel the new setting and go to step 8.

Quitting the operation

Step8 The balance displays the next menu (**LP Fnc**) of the function table.

Press the **CAL** **key** to exit the clock and calendar function and return to the weighing mode.

Note Do not enter invalid values such as a non-existing date when setting the time and date.

When the clock backup battery has been depleted, the balance displays **rtc PF**. Under this condition, press any key and set the time and date. The dead battery only affects the clock and calendar function. Even so, the function works normally as long as the AC adapter is connected to the balance.

10-5 Comparator Function

The comparison of comparators can select 3-steps or 5-steps ($\overline{CP} Fnc$, $\overline{CP} -t$), and it is set to 3-steps at the factory setting.

When 3-step comparator is set, the results of the comparison are indicated by HI OK LO on the display.

When 5-step comparator is set, HH is indicated by HI blinking and LL by LO blinking.

By using GXA-04, it is possible to output the comparison result at the contact point.

There are three types of scope that can be selected as follows.

- No comparison
- Comparison when the weight data is stable or overloaded
- Continuous comparison

The conditions for comparing near zero are in six levels from "including near zero" to " ± 100 digits".

"Upper limit value and lower limit value" and "reference value and tolerance range" are the comparison standards.

"Digital input" and "Input by sample load" are the for each value input methods for each value.

Refer to the function setting $\overline{CP} Fnc$.

By setting the function setting $\overline{CP} bEEP$, it is also possible to sound an internal buzzer depending on the result of the comparison.

3-step comparison result

Weighing value Threshold value ↑ Judgment formula		3-step comparison - display			
		Judgment result	Lit display	Blinking display	Buzzer control
Upper limit	Upper limit value < Weighing value	HI	HI	/	bEP HI
Lower limit	Lower limit value \leq Weighing value \leq Upper limit value	OK	OK		bEP OK
		LO	LO		bEP LO

5-step comparison result

Weighing value Threshold value ↑ Judgment formula		5-step comparison - display			
		Judgment result	Lit display	Blinking display	Buzzer control
Second upper limit	2nd Upper limit value < Weighing value	HH	/	HI	bEP HH
Upper limit	Upper limit value < Weighing value \leq 2nd Upper limit value	HI		HI	bEP HI
Lower limit	Lower limit value \leq Weighing value \leq Upper limit value	OK		OK	bEP OK
Second lower limit	2nd Lower limit value \leq Weighing value < Lower limit value	LO	LO	bEP LO	
	Weighing value < 2nd Lower limit value	LL	/	LO	bEP LL

Note

- In the flow measurement, the comparator function is used with the total weight.

Example1 (Comparison when the weighing data is stable or overloaded, excluding "near zero" ± 20 digit, upper limit and lower limit.)

Selecting a comparator mode

- Step1 Press and hold the **SAMPLE** key until **bASFnC** of the function table is displayed, then release the key.
- Step2 Press the **SAMPLE** key several times to display **[P Fnc]**.
- Step3 Press the **PRINT** key.
- Step4 Press the **RE-ZERO** key several times to display **[P-ALL]** ("2" always compare).
- Step5 Press the **SAMPLE** key several times to display **[P-Z]**.
- Step6 Press the **RE-ZERO** key several times to display **[P-Z EX 20d]** ("3" ± 20 digit is not compared.)
- Step7 Press the **SAMPLE** key several times to move to **[P in]**.
- Step8 Press the **RE-ZERO** key several times to display **[P in rFE DIG]**. ("2" reference value is set. digital input)
- Step9 Press the **PRINT** key to store the selected mode.

Entering the upper and lower limit values

- Step10 With **[P VALUE]** displayed, press the **PRINT** key to display **[P rEF]**.
- Step11 Press the **PRINT** key. The current setting of the upper limit value is displayed with all of the digits blinking. When the current setting is not to be changed, press the **PRINT** or **CAL** key to go to step 12.

When the current setting is to be changed, press the **RE-ZERO** key and store the following keys.

- SAMPLE** key To select the digit to change the value.
- RE-ZERO** key To change the value of the digit selected.
- MODE** key To switch the polarity.
- PRINT** key To store the new setting and go to step10.
- CAL** key To cancel the new setting and go to step 10.

- Step12 When **[P LME]** is displayed, pressing the **PRINT** key will display the currently set value. If changing the setting value, it can be registered the tolerance value with the following keys.

For tolerance value, enter the value with the reference value set to 100%.

- SAMPLE** key Move the blinking digit.
- RE-ZERO (+)**key Change the value of the blinking digit.
- MODE (-)**key Change the value of the blinking digit.
- PRINT** key Register and go to Step13
- CAL** key Cancel and go to Step13

- Step13 Press the **CAL** key to return to the weighing display.

Example 2

(Continuous comparison, including "near zero", reference value and tolerance value.)

Selecting a comparator mode

- Step1 Press and hold the **SAMPLE** key until **bASFnC** of the function table is displayed, then release the key.
- Step2 Press the **SAMPLE** key several times to display **[CP Fnc]**.
- Step3 Press the **PRINT** key.
- Step4 Press the **RE-ZERO** key several times to display **[CP Z TRb]**. ("I" compared when stable and over)
- Step5 Press the **SAMPLE** key several times to display **[CP Z]**.
- Step6 Press the **RE-ZERO** key several times to display **[CP Z IN]**. ("[]" near zero is also compared.)
- Step7 Press the **SAMPLE** key several times to display **[CP in]**.
- Step8 Press the **RE-ZERO** key several times to display **[CP in H/LWT]**. ("I" upper-lower limit is set. Input by loaded.)
- Step9 Press the **PRINT** key to store the new setting.

Entering the reference and tolerance values

- Step10 When **[CP VALUE]** is displayed, press the **PRINT** key. **[CP Hi]** will be displayed.
- Step11 When **[CP Hi]** is displayed, press the **PRINT** key to check the currently set value (all blinking). Press the **RE-ZERO** key to enter the load input mode.
- Step12 Press the **RE-ZERO** key, **0.00g** is displayed. Place a sample of the weight of the upper limit on the balance and press the **PRINT** key. (Register the upper limit value.)
- Step13 When finished, **[CP Lo]** is displayed. (Replace a sample of the weight of the upper limit from the balance.)
- Step14 When **[CP Lo]** is displayed, press the **PRINT** key to check the currently set value (all blinking). Press the **RE-ZERO** key to enter the load input mode.
- Step15 Press the **RE-ZERO** key, **0.00g** is displayed. Place a sample of the weight of the lower limit on the balance and press the **PRINT** key. (Register the lower limit value.)
- Step16 When finished, **[CP Hi]** is displayed. (Replace a sample of the weight of the lower limit from the balance.)
- Press the **CAL** key twice to return to the weighing display.

Sound the built-in buzzer corresponding to the comparison result.

- Step1 Press and hold the **SAMPLE** key until **bASFnC** of the function table is displayed.
- Step2 Press the **SAMPLE** key several times to display **[CP BEEP]**.
- Step3 Press the **PRINT** key.
- Step4 Press the **SAMPLE** key to set the buzzer sound setting of the comparison result "ON/OFF".

When 3-step comparator is set, the display can be selected from the following 3 kinds

[bEP Hi] **[bEP ok]** **[bEP Lo]** .

When 5-step comparator is set, the display can be selected from the following 5 kinds

[bEP HH] **[bEP Hi]** **[bEP ok]** **[bEP Lo]** **[bEP LL]** .

SAMPLE key Select the comparison result.

RE-ZERO key Set the buzzer sound setting of the comparison result "ON/OFF"

PRINT key Store the setting.

CAL key Cancel and return to the weighing display.

11.ID Number And GLP Report

- The ID number is used to identify the balance when Good Laboratory Practice (GLP) or Good Manufacturing Practice (GMP) is used.
- The data output compatible with "GLP/GMP" can be output to a personal computer or printer using the RS-232C serial interface.
- The GLP / GMP compliant report includes the balance manufacturer, model, serial number, ID number, date, time and space for signature. It includes the results and using mass for calibration or calibration test data.
- The balance can output the following reports for GLP/ GMP.
 - "Calibration report" of the calibration, using the internal mass (Calibration due to changes in temperature and one-touch calibration.)
 - "Calibration report" of the calibration, using an external weight.
 - "Calibration test report" of the calibration test, using an external weight.
 - "Title block" and "End block" for the weighing data.
- Calibration and calibration test data can be stored in memory to output several reports at the same time. Refer to "12. Data Memory" for details.
- The ID number is used to identify the balance when the balance is used for maintenance management.
- The ID number is maintained in non-volatile memory even if the AC adapter is removed.
- For details on confirming and setting the time and date. Refer to "10.4. Clock and Calendar Function".

11-1 Setting The ID Number

Step1 Press and hold the **SAMPLE** key until **bASFnC** of the function table is displayed, then release the key.

Step2 Press the **SAMPLE** key several times to display **id**.

Step3 Press the **PRINT** key. Set the ID number using the following keys.

SAMPLE key..... To select the digit to change the value.

RE-ZEROkey, **MODE** key To set the character of the digit selected.

..... Refer to the display character set shown below.


PRINT key..... To store the new ID number and display **PASSwd**.

CAL key To cancel the new ID number and display **PASSwd**.

Step4 With **PASSwd** displayed, press the **CAL** key to return to the weighing mode.

Note The display segment of the balance is divided into 4types.

For each segment display, refer to the "Display correspondence table" in the next page.



Display correspondence table

11 Segment

0	1	2	3	4	5	6	7	8	9	-	␣	A	B	C	D	E	F	G	H	I	J	K	L	M	N	O	P	Q	R	S	T	U	V	W	X	Y	Z
0	1	2	3	4	5	6	7	8	9	-		A	B	C	D	E	F	G	H	I	J	K	L	M	N	O	P	Q	R	S	T	U	V	W	X	Y	Z

↳ Space

7 Segment

0	1	2	3	4	5	6	7	8	9	-	␣	A	B	C	D	E	F	G	H	I	J	K	L	M	N	O	P	Q	R	S	T	U	V	W	X	Y	Z
0	1	2	3	4	5	6	7	8	9	-		A	b	c	d	E	F	G	H	I	J	K	L	M	N	O	P	Q	r	S	T	U	V	W	X	Y	Z

↳ Space

14 Segment

0	1	2	3	4	5	6	7	8	9	-	␣	A	B	C	D	E	F	G	H	I	J	K	L	M	N	O	P	Q	R	S	T	U	V	W	X	Y	Z
0	1	2	3	4	5	6	7	8	9	-		A	B	C	D	E	F	G	H	I	J	K	L	M	N	O	P	Q	R	S	T	U	V	W	X	Y	Z

↳ Space

15 Segment

0	1	2	3	4	5	6	7	8	9	-	␣	A	B	C	D	E	F	G	H	I	J	K	L	M	N	O	P	Q	R	S	T	U	V	W	X	Y	Z
0	1	2	3	4	5	6	7	8	9	-		A	B	C	D	E	F	G	H	I	J	K	L	M	N	O	P	Q	R	S	T	U	V	W	X	Y	Z

↳ Space

11-2 GLP Report

Set the function setting to " *Info I*" for output of GLP / GNP data.

- Output to a mini-printer "AD-8121B".
 - Set the "GLP report (*Info*)" parameter to " *I*" .
 - Turn on the DIP switch No. 2 of the printer and change the print mode to MODE 3.
 - Output to a mini-printer "AD-8126".
 - Set the "GLP report (*Info*)" parameter to " *2*" .
- Output to a multi-printer "AD-8127".
 - Set the "GLP report (*Info*)" parameter to " *2*" .
 - Select the print mode to **DUMP** .
- Output to a PC
 - Set the "GLP report (*Info*)" parameter to " *I*" .

Note

If the time and date are not correct set the correct time and date in "Clock (*CLRD*)" of the function table.

Calibration report using the internal mass

This is the GLP report when the balance is calibrated using the internal mass.

Setting of *Info I*

Printer format

```

      A & D
MODEL  GX-10002A
S/N    123456789
ID     LAB-0123
DATE   2017/12/31
TIME   12:34:56
CALIBRATED (INT.)
REMARKS

SIGNATURE
-----
  
```

Setting of *Info I*

PC format

```

_____ A-&-D<TERM>
MODEL_ GX-10002A<TERM>
S/N___ 123456789<TERM>
ID____ LAB-0123<TERM>
DATE__ 2017/12/31<TERM>
TIME___ 12:34:56<TERM>
CALIBRATED (INT.)<TERM>
REMARKS<TERM>
<TERM>
<TERM>
SIGNATURE<TERM>
<TERM>
<TERM>
-----<TERM>
<TERM>
<TERM>
  
```

_ Space, ASCII 20h.
 <TERM> Terminator, CR LF or CR
 CR Carriage return, ASCII 0Dh.
 LF Line feed, ASCII 0Ah.

Calibration report using an external weight

This is the GLP report when the balance is calibrated using the external weight.

Setting of *inF0* |

Printer format

```

                A & D
MODEL    GX-10002A
S/N      123456789
ID       LAB-0123
DATE     2017/12/31
TIME     12:34:56
CALIBRATED<EXT.>
CAL.WEIGHT
          +100000.00 g
REMARKS

SIGNATURE
-----
    
```

Setting of *inF0* |

PC format

```

                A & D <TERM>
MODEL _GX-10002A <TERM>
S/N ___123456789 <TERM>
ID ___  LAB-0123 <TERM>
DATE _2017/12/31 <TERM>
TIME ___ <TERM>
CALIBRATED (EXT.) <TERM>
CAL. WEIGHT <TERM>
    ___+10000.00 __g <TERM>
REMARKS <TERM>
<TERM>
SIGNATURE <TERM>
<TERM>
----- <TERM>
<TERM>
<TERM>
    
```

— Space, ASCII 20h

<TERM>Terminator, CR LF or CR*

CR Carriage return, ASCII 0Dh*

LF Line feed, ASCII 0Ah*

Calibration test report using an external weight

This is the GLP report when checking the weighing accuracy of the balance with the external weight.

(Adjustment is not performed)

Setting of *inF0* |

Printer format

```

                A & D
MODEL    GX-10002A
S/N      123456789
ID       LAB-0123
DATE     2017/12/31
TIME     12:34:56
CAL. TEST<EXT.>
ACTUAL
          0.00 g
          +99999.95 g
TARGET
          +100000.00 g
REMARKS

SIGNATURE
-----
    
```

Setting of *inF0* |

PC format

```

                A & D <TERM>
MODEL _GX-10002A <TERM>
S/N ___123456789 <TERM>
ID ___  LAB-0123 <TERM>
DATE _2017/12/31 <TERM>
TIME ___12:34:56 <TERM>
CAL. TEST (EXT.) <TERM>
ACTUAL <TERM>
    ___+0.00 __g <TERM>
    ___+99999.95 __g <TERM>
TARGET <TERM>
    ___+100000.00 __g <TERM>
REMARKS <TERM>
<TERM>
SIGNATURE <TERM>
<TERM>
<TERM>
    
```

— Space, ASCII 20h

<TERM>Terminator, CR*LF or CR

CR Carriage return, ASCII 0Dh

LF Line feed, ASCII 0Ah

Heading and ending output

Application / Operation

As a method of managing weighing values, add "Heading" and "End" parts before and after the weighing value.

By pressing and holding the **PRINT** key, "Heading" and "End" are output in turn.

Note

If the data memory function is used (except when *DATA 0*), heading and end cannot be output..

Key output method

Step1 While displaying the weighing value, hold down the **PRINT** key and display **Start** to output "Heading".

Step2 Output the weighing value. The output method depends on the setting of the data output mode.

Step3 Press and hold the **PRINT** key to display **End**, "End" is output.

Setting of *info 1*

Printer format

(Internal setting *TYPE 1*)

```

                A & D
MODEL    GX-10002A
S/N      123456789
ID       LAB-0123
DATE    2017/12/31
START
TIME     12:34:56

WT      +12.3456 g
WT      +12.3461 g
WT      +12.3462 g
WT
  
```

```

WT      +12.3461 g
WT      +12.3453 g
WT      +12.3471 g
WT      +12.3464 g
  
```

```

END
TIME     12:45:56
REMARKS
  
```

```

SIGNATURE
-----
  
```

Setting of *info 1*

PC format

(Internal *TYPE 1*)

```

                A & D<TERM>
MODEL _GX-10002A<TERM>
S/N ___123456789<TERM>
ID ____LAB-0123<TERM>
DATE ___2017/12/31<TERM>
START<TERM>
TIME____12:34:56<TERM>
<TERM>
  
```

```

WT ___+12.3456 __g<TERM>
WT ___+12.3461 __g<TERM>
WT ___+12.3462 __g<TERM>
  
```

```


WT ___+12.3451 __g<TERM>
WT  +12.3453  g<TERM>
WT  +12.3471  g<TERM>
WT  +12.3464  g<TERM>
  
```

```

<TERM>
END<TERM>
TIME  12:45:56<TERM>
  
```

```

REMARKS<TERM>
<TERM>
<TERM>
SIGNATURE<TERM>
<TERM>
<TERM>
  
```



- ___ Space, ASCII 20h
- <TERM> terminator, CR LF or CR
- CR Carriage return, ASCII 0Dh
- LF Line feed, ASCII 0Ah

12.Data Memory

Data memory is a function to store weighing data and calibration data in memory. The data stored in memory are available for outputting at one time to a printer or personal computer.

The following five types of data can be stored.

Unit mass (Counting mode)	Up to 50 sets
Weighing value	Up to 200 sets
Calibration report Internal calibration External calibration Calibration test report Internal test calibration External test calibration	Last 50 sets

12-1 Data Memory For Weighing Data

Features


- It is not necessary to connect the printer or personal computer to the balance continually, because the balance stores the weighing data in memory.
- The data in memory can be displayed on the balance for confirmation.
- Data (ID number, data number, time and date) to be added to the output data can be selected in the function setting.
- The balance can store 200 sets of weighing data in memory (if time and date are added, the balance can store 100 sets).
 - ※ For the unit mass storage method, refer to "5-2 Counting Mode".

Storing the weighing data

Note

1. Set the "Data memory (dAtA)" parameter to " dAtA 2". Refer to "10. Function Table".
2. Specify the "Time/Date output (S-t-d)" parameter whether time and date is to added or not.
3. The storing mode depends on the "Data output mode (Pr-t)" parameter setting. When set to Pr-t3 (stream mode), data may not be stored correctly.
 - ※ It is also possible to change the time / date setting after storing the weighing value.

Display and symbol



Caution

- When weighing data is being stored in memory, the data is output simultaneously using RS-232C interface or USB.
- "FUL" indicates that memory is full or the memory capacity has been reached. More data cannot be stored unless the memory data is deleted.
- Automatic self calibration can not be used while the interval memory mode is active.
- Statistic calculation function can not be used when the data memory function is active.

Setting the function table

Parameter settings for each output mode are as follows:

Mode \ Item	Data output mode	Auto print polarity, difference	Data memory function	Interval time
Key mode	<i>Prt 0</i>	Not used	<i>dAtA 2</i>	Not used
Auto print mode A	<i>Prt 1</i>	<i>AP-A 0 ~ 2</i>	<i>dAtA 2</i>	
Auto print mode B	<i>Prt 2</i>	<i>AP-b 0 ~ 2</i>	<i>dAtA 2</i>	
Key mode B (immediate)	<i>Prt 4</i>	Not used	<i>dAtA 2</i>	
Key mode C (stable)	<i>Prt 5</i>		<i>dAtA 2</i>	
Interval output mode	<i>Prt 6</i>		<i>dAtA 2</i>	

Parameter settings for Data number, ID number, Time and Date

Data number	No	<i>d-no "0"</i>	Time and date	No	<i>S-t d 0</i>	Up to 200 pieces
	Yes	<i>d-no "1"</i>		Time only	<i>S-t d 1</i>	
ID number	No	<i>S-id "0"</i>		Date only	<i>S-t d 2</i>	
	Yes	<i>S-id "1"</i>		Both	<i>S-t d 3</i>	

Enabling the data memory function

- 1 Press and hold the **SAMPLE** key until **bASFnC** is displayed, then release the key.
- 2 Press the **SAMPLE** key several times to display **dout**.
- 3 Press the **PRINT** key.
- 4 Press the **SAMPLE** key three times to display **dAtA**.
- 5 Press the **RE-ZERO** key to display **dAtA wt/RL**.
- 6 Press the **PRINT** key to store the setting.
- 7 Press the **CAL** key to return to the weighing mode.

Recalling the memory data

Confirm that the "Data memory (*dAtA*)" parameter is set to "*dAtA 2*".

1. Press and hold the **PRINT** key until **RECALL** is displayed, then release the key.

2. Press the **PRINT** key to enter the memory recall mode. The type of data appears in the upper left of the display as shown to the right "-d-" or "d-t". Recall the data in memory using the following keys.

RE-ZERO key To proceed to the next data set.

MODE key To go back to the previous data set.

PRINT key To transmit the current data using the RS-232C or USB.

With **SAMPLE** held down, press the **CAL** key

To delete the current data.

CAL key To exit the memory recall mode.

3. Press the **CAL** key to return to the weighing mode.

※ It is also possible to change the time / date output setting after storing the weighing value.

Left of the display

-d-

When setting without clock / date

or

d-t

When setting with clock / date

Transmitting all memory data at one time

Confirm that the "Serial interface (S_rF)" parameters are set properly.

Refer to "10. Function Table" and "Communication Manual " on the A&D website.

1. Press and hold the **PRINT** key until **RECALL** is displayed, then release the key.
2. Press the **SAMPLE** key to display **out**.
3. Press the **PRINT** key to display **out** with "No" blinking.
4. Press the **RE-ZERO** key to display **out** with "No" blinking.
5. Press the **PRINT** key to transmit all data using the RS-232C , USB.
6. The balance displays **CLEAR** when all data is transmitted.
Press the **CAL** key to return to the weighing mode.

Deleting all memory data at one time

1. Press and hold the **PRINT** key until **RECALL** is displayed, then release the key.
2. Press the **SAMPLE** key several times to display **CLEAR**.
3. Press the **PRINT** key to display **CLEAR** with "No" blinking.
4. Press the **RE-ZERO** key to display **CLEAR** with "No" blinking.
5. Press the **PRINT** key to delete all data
6. The balance displays **End** and returns to the weighing mode.


12-2 Data Memory For Calibration And Calibration Test

Characteristic

- Calibration data (when and how it is performed) and calibration test data can be stored in memory.
- All the data in memory is available to be output at one time to a printer or personal computer.

Up to 50 data sets of the latest calibration or calibration test can be stored.

※ When the memory capacity has been reached to 50, "FUL" ↔ "CAL" illuminates in order in the upper left of the display as shown below.



Storing the calibration and calibration test data

1. Set the "Data memory (dARA)" parameter to " dARA 2". Refer to "10 Function Table".
2. With the settings above, each time calibration or calibration test is performed, the data is stored automatically.

Transmitting the memory data

Step1 Press and hold the **CAL** key during weighing display. When a **CAL H.5** displayed, release your finger from the key to display **out**.

Step2 Press the **PRINT** key to display **out No**.

Step3 Change the **No** / **Go** with the **RE-ZERO** key.

Display the **out Go**.

Step4 Press the **PRINT** key to start output at one time while **out Go** is displayed.


The output format conforms to "GLP output".

Step5 When output at one time is completed, **CLEAR** displays after **End** is displayed.

Step6 If the saved history is deleted all at once, please proceed to How to delete history. To return to the weighing value, press the **CAL** key.


* If the **FULL** ↔ **CAL** indicators blink in turn during weighing display, 50 instance of data are stored.

If history is saved history in this state, old data will be overwritten. Optionally delete the saved data.



Deleting data stored in memory

1. Press and hold the **CAL** key until **CAL H.5** is displayed, then release the key. **out** is displayed.
2. Press the **SAMPLE** key to display **CLEAR**.
3. Press the **PRINT** key to display **CLEAR No.**
4. Press the **RE-ZERO** key to change **No.** / **Go.**
Display **CLEAR Go.**
5. Press **PRINT** while **CLEAR Go.** is displayed, output at once is started.
6. When the balance displays **End** and returns to the weighing mode.



13. Statistical Calculation Mode

The statistical calculation mode statistically calculates the weight data, and displays or outputs the results. To use the statistical calculation mode, set the "Application function (APF)" parameter of "Application (AP Fnc)" in the function table to "2", as described below.

Statistical items available are number of data, sum, maximum, minimum, range (maximum-minimum), average, standard deviation and coefficient of variation. What statistical items to output can be selected from the four modes in the function table (5t APF).

- The wrong data input can be canceled by the key operation, if immediately after the input.
- Turning the balance off will delete the statistical data.

- The standard deviation and coefficient of variation are obtained by the equation below:

$$\text{Standard deviation} = \sqrt{\frac{N \cdot \sum (X_i)^2 - (\sum X_i)^2}{N \cdot (N-1)}} \quad \text{where } X_i \text{ is the } i\text{-th weight data,}$$

N is number of data.

$$\text{Coefficient of variation (CV)} = \frac{\text{Standard deviation}}{\text{Average}} \times 100 (\%)$$

$$\text{Relative error of maximum value} = \frac{\text{Maximum value} - \text{Average}}{\text{Average}} \times 100 (\%)$$

$$\text{Relative error of minimum value} = \frac{\text{Minimum value} - \text{Average}}{\text{Average}} \times 100 (\%)$$

- When there is data with a minimum display digit off, the calculation result is displayed with the minimum display digit off. (Minimum display digit is rounded off.)

13-1 How To Use The Statistic Calculation

(1) Getting Started

Switching to the Statistical Function Mode (Changing The Function Table)


1. Press and hold the **SAMPLE** key until **bAS5Fnc** of the function table is displayed, then release the key.
2. Press the **SAMPLE** key several times to display **AP Fnc**.
3. Press the **PRINT** key to display **APF 5t**.
4. Press the **RE-ZERO** key several times to display **APF 5t**.

To select statistical items to output, go to step 5.


To store the statistical function mode setting, go to step 7.

To disable the statistical calculation mode, press the

RE-ZERO key to select **APF 5t**.




Selecting the statistical items to output

- Press the **SAMPLE** key to display .
- Press the **RE-ZERO** key to select the output items. In the example, output the number of data, sum, maximum, minimum, range (maximum-minimum) and average are selected.


Parameter	Description
0	Number of data, sum
1	Number of data, sum Maximum, minimum, range (maximum – minimum), average
2	Number of data, sum Maximum, minimum, range (maximum – minimum), average, Standard deviation, coefficient of variation
3	Number of data, sum Maximum, minimum, range (maximum – minimum), average, Standard deviation, coefficient of variation Relative error of maximum value, relative error of minimum value

- Press the **PRINT** key to store the setting.
- Press the **CAL** key to return to the weighing mode.



Selecting the unit

- Press the **MODE** key to select the unit to be used for the statistical calculation mode. In the example shown at the right, gram (g) is selected.



Notes

Selecting the unit using the **MODE** key is not available after the data is entered. In this case, clear the data as described on page 67 and select the unit using the **MODE** key.

When the unit used for the statistical calculation mode is to be enabled upon power-on, select the unit in “Unit (Unit)” of the function table beforehand.

(2) Using The Statistical Calculation Mode

Entering data for statistical calculation

Use the following keys to operate the statistical calculation mode.

MODE key When the data is entered, moves between the displaying items (weighing mode, statistical results and data operation) each time the key is pressed.
When no data has been entered, selects the unit.

SAMPLE key Turns the minimum weighing value on or off, in the weighing mode.


RE-ZERO key Sets the display to zero in the weighing mode.

PRINT key Outputs the data number and the weight data and includes the weight data to statistical calculation in the weighing mode. (Output is not in the data format specified in the function table because of the data number added.)

Outputs the statistical results while the statistical results are displayed. (Output is not in the data format specified in the function table.)


CAL key Returns to the weighing mode.

- 1 Press the **RE-ZERO** key to set the display to zero.
- 2 Place the sample on the weighing pan and wait for the stabilization indicator to turn on.
- 3 Press the **PRINT** key to add the data displayed to statistical calculation. The number of data on the upper left of the display increases by 1.
- 4 Repeat steps 1 to 3 for each weighing.



Outputting the statistical results

- 5 Each time the **MODE** key is pressed, the display changes: the results as selected in “Statistical function mode output items (SEAF)”, and **CLEAR**, **CANCEL**.



Notes

When the number of data is 1, the coefficient of variation is displayed as **-----**.

When the average is 0, the coefficient of variation is displayed as **-----**.

Statistical items are indicated on the upper left of the display using the following symbols.

Symbol	Statistical item
$\Sigma \bar{x}$	Sum
\bar{x}_{MAX}	Maximum
\bar{x}_{MIN}	Minimum
r	Range (Maximum – minimum)
\bar{x}	Average
Sd	Standard deviation
Cv	Coefficient of variation
$\bar{x}_{MAX}\%$	Relative error of maximum value
$\bar{x}_{MIN}\%$	Relative error of minimum value


Output example

Function table parameter	
N	10
SUM	1000.00g
MAX	105.00g
MIN	95.00g
R	10.00g
AVE	100.00g
SD	2.800g
CV	2.8 %
MAX%	5.0 %
MIN%	5.0 %

(3) Deleting the latest data



When the wrong data is entered, it can be deleted and excluded from statistical calculation. Only the latest data can be deleted.


- In the weighing mode, press the **MODE** key to display **CANCEL**.
- Press the **PRINT** key to display **CANCEL**.
- Press the **RE-ZERO** key to display **CANCEL**.
- Press the **PRINT** key to delete the latest data and exclude it from statistical calculation.
The number of data decreases by 1 when the balance returns to the weighing mode.



Clearing the statistical data

All the statistical data will be deleted and the number of data will be 0 (zero).

- 1 In the weighing mode, press the **MODE** key, to display **CLEAR**.
- 2 Press the **PRINT** key to display **CLEAR** .
- 3 Press the **RE-ZERO** key to display **CLEAR** .
- 4 Press the **PRINT** key to delete the statistical data. The number of data becomes 0 (zero) when the balance returns to the weighing mode.



13-2 Statistical Calculation Mode (Example Of Use)

Here, as an example of use of the statistical calculation mode, mixing of the multiple formulae such as medicine is described. The mixing process is recorded using the balance and the printer.

In the example, the GX-303A and the AD-8126 or AD-8127 are connected using the RS-232C serial interface.

(1) Getting Started

Changing the function table

- Changes
 - To enable the statistical calculation mode
 - To enable “Zero after output”

Enabling the statistical calculation mode

- 1 Enter the function table menu.

Press and hold the **[SAMPLE]** key until **bASFnC** of the function table is displayed, then release the key.

- 2 Select the application function.

Press the **[SAMPLE]** key several times to display **APFnC**.

Then, press the **[PRINT]** key to display **•APF Np-n**.

- 3 Change the application function parameter to “2”.

Press the **[RE-ZERO]** key to display **2APF S-zit**. Press the **[PRINT]** key to confirm the change. After **End**, **MWFnC** is displayed.

Enabling “Zero after output”

- 4 Select “Zero after output”.

Press the **[SAMPLE]** key several times to display **dout**. Then,

press the **[PRINT]** key to display **•Pr-t KE3**, and press the

[SAMPLE] key several times to display **•Pr-d JFF**.


- 5 Enable “Zero after output”.

Press the **[RE-ZERO]** key to display **Pr-d Jv**. Then, press

the **[PRINT]** key to confirm the change. After **End**, **SIF** is displayed.

Returning to the weighing mode

- 6 Press the **[CAL]** key to return to the weighing mode.




(2) Using The Statistical Calculation Mode

- 1 Press the **RE-ZERO** key to set the display to zero.
- 2 Place a container on the weighing pan.
Press the **PRINT** key to cancel the weight (tare). The balance displays **0.000 g**. (Storing the tare value)
The tare value data is output when the peripheral output equipment is connected.
- 3 Weigh formula 1 and press the **PRINT** key. The balance displays **0.000 g**. (Storing the weight value of formula 1)
The weight value data is output when the peripheral output equipment is connected.
- 4 Weigh formula 2 and press the **PRINT** key. The balance displays **0.000 g**. (Storing the weight value of formula 2)
- 5 When there are some more formulae to be added, repeat step 4.
- 6 After mixing is complete, press the **MODE** key to display the statistical results.
- 7 Press the **PRINT** key to output the number of data saved including the tare value and the total weight.

Output example

No. 1		
ST, +0005.637 g	-----	Tare value
No. 2		
ST, +0001.992 g	-----	Formula 1
No. 3		
ST, +0007.780 g	-----	Formula 2
N	3	
SUM	+15.409 g	----- Total weight



14. Flow Measurement

The balance has a "flow mode" that calculates the amount of change in the weighing value per hour. To use the statistical calculation mode, set the "Application function (APF)" parameter of "Application (APP Fnc)" in the function table to "3", as described below.

Calculating the flow rate from the weight output of the balance may change the number of weighing values per unit time, so there is a possibility that the result will not be correct.


The flow mode is calculated at the accurate timing at which the weighing value is updated in the balance, so it is possible to calculate the correct flow rate value.

If the flow unit is set to ml/*, density can be registered. The maximum number of registrations is 10, and if density is set in advance, it can be selected according to the measurement sample.


Flow calculation formula

$$Q = \frac{W - W'}{Ct}$$

Q : Flow rate Ct : Calculation time
W : Current calculated value W' : Weight value before Ct



Ct = 1 second
Calculate instantaneous flow rate



Ct = 5 minutes
With the amount of change in the measured value for 5 seconds, calculate flow rate

The longer the Ct is set, the more stably the flow rate value can be calculated.

Note

In flow mode, the display update timing during flow display changes according to the calculation time Ct.

The relationship between calculation time Ct and flow update timing is as follows.

Calculation Time Ct	Display update	Calculation Time Ct	Display update	Calculation Time Ct	Display update
1 sec.	1 sec.	30 sec.	1 sec.	20 min.	10 sec.
2 sec.	1 sec.	1 min.	1 sec.	30 min.	15 sec.
5 sec.	1 sec.	2 min.	1 sec.	1 hour	30 sec.
10 sec.	1 sec.	5 min.	3 sec.		
20 sec.	1 sec.	10 min.	5 sec.		

- The flow rate value is displayed as "0" until the set calculation time Ct has elapsed.
- The recorded weighing data can be initialized by turning off the power or by pressing **RE-ZERO** key.
- In flow mode, the unit is g and flow unit only. Other units cannot be displayed.
- Automatic environment setting / self-diagnosis function cannot be used in flow mode.
- "Heading" and "End" of the GLP report function cannot be output in flow mode.
- The data memory function cannot be used in flow mode.

14-1 How To Use Flow Measurement

Enable flow rate measurement


Switch flow rate measurement (Change the Function Table)

- 1 Press and hold the **SAMPLE** key until **bASFnC** of the function table is displayed, then release the key.
- 2 Press the **SAMPLE** key several times to display **RP Fnc**.
- 3 Press the **PRINT** key to display **0 APF N/A**.
- 4 Press the **RE-ZERO** key several times to display **3 APF Fnd**.

If you want to change the flow rate unit, go to "5".

If you want store the setting, go to "7".

If you want to cancel the flow function, press the **RE-ZERO** key and return to **0 APF N/A**.



Setting of flow rate unit

- 5 Press the **SAMPLE** key to display **0 Fnd Unit**.
- 6 Press the **RE-ZERO** key to set the setting value.



Parameter	Contents
0	g / s (gram/sec)
1	g / m (gram/min)
2	g / h (gram/hour)
3	mL / s (milliliter/sec)
4	mL / m (milliliter/min)
5	mL / h (milliliter/hour)

- 7 Press the **PRINT** key to store.
- 8 Press the **CAL** key to return to the calculating display.

Change display

After returning to the weighing value display after setting to flow mode, the unit is "g" with the F_{rd} or F^{**} indicator on.

Use the $\boxed{\text{MODE}}$ key to switch between flow rate display and "g" display. By switching, the total amount and flow rate can be checked.




How to set calculation time

In weighing display, press and hold the $\boxed{\text{MODE}}$ key to display $\boxed{Et^{***}}$.

Calculation time can be changed by following key operation.

The setting range is 1 second to 1 hour.




- $\boxed{\text{RE-ZERO}}$ (+) key Change calculation time
- $\boxed{\text{MODE}}$ (-) key Change calculation time
- $\boxed{\text{PRINT}}$ key Store setting value
- If the flow rate unit is g / *,
the display will return to weighing display.
- When the flow rate unit is mL / *,
the display goes to density setting display.

$\boxed{\text{CAL}}$ key It returns to weighing display without storing the set value.

When the calculation time automatic setting $\boxed{Et Auto}$ is "1", the resolution of the flow rate calculation result can be set.

The larger the number, the more precision will be calculated.



- $\boxed{\text{RE-ZERO}}$ (+) key Change the resolution
- $\boxed{\text{PRINT}}$ (-) key Store setting value
- If the flow rate unit is g / *,
the display will return to weighing display.
- When the flow rate unit is mL / *,
the display goes to density setting display.

$\boxed{\text{CAL}}$ key It returns to weighing display without storing the set value.


How to set the density

When the setting value of function setting $F_{rd}, Unit$ is 3, 4, 5, after setting the calculation time, go to density setting display.

Density can be changed by following key operation.

The setting range is 0.0001g/cm³ to 9.9999g/cm³.


- $\boxed{\text{RE-ZERO}}$ (+) key Change the number of the blinking digit
- $\boxed{\text{MODE}}$ (-) key Change the number of the blinking digit
- $\boxed{\text{SAMPLE}}$ key Move the blinking digit
- $\boxed{\text{PRINT}}$ key The set value is the display returns to weighing display.
- $\boxed{\text{CAL}}$ key The display returns to weighing display without storing the set value.



Method of reading density number

When flow unit is mL/*, up to 10 densities can be registered.

To register a new density, read the unconfigured density number and then register according to the procedure of the setting method of calculation time.



Continuing to hold down the **PRINT** key in weighing display displays $d^{*}.****$.

Blinking F^{**} is the current density number and $d^{*}.****$ is the set density value.

The density number can be changed by following key operation

The setting range is F01~F10.

- RE-ZERO**(+) key Change density number.
- MODE**(-) key Change density number.
- PRINT** key Read the density of the selected density number and return to weighing display.
- CAL** key Return to the weighing display without reading the density of the selected density number.

14-2 Flow Measurement Setting

Flow value range by models

Model	Flow value range
0.001g model	0.01~100
0.01g model	0.1~1000
0.1g model	1~5000

Depending on the sample to measure the flow rate, the above values may not apply.

Estimated flow rate value and calculating time (Ct) depending on your model

Calculation time automatic setting mode

By setting "Calculation time automatic mode" of Function table "Application Function $AP F_{nc}$ " to "1", the optimal Ct is automatically set for the flow rate being measured. So, it is unnecessary to manually set Ct.

In the case of using 0.1g model

Flow rate (mL/min)	Response priority ← Ct setting → Accuracy priority									
	1sec	2sec	5sec	10sec	20sec	30sec	1min	2min	5min	10min
0.01										
0.02										
0.05										
0.1										
0.2										
0.5										
1										
2										
5										
10										
20										
50										
100										

Example) When measuring the flow rate of 0.02 ml/min

Set the calculation time in 5minutes or more

In the case of using 0.01g model

Flow rate (mL/min)	Response priority ← Ct setting → Accuracy priority									
	1sec	2sec	5sec	10sec	20sec	30sec	1min	2min	5min	10min
0.1										
0.2										
0.5										
1										
2										
5										
10										
20										
50										
100										
200										
500										
1000										
200										
500										
1000										

Example) When measuring the flow rate of 20 ml/min

Set the calculation time in 5 seconds to 30 seconds

In the case of using 0.1g model

Flow rate (mL/min)	Response priority ← Ct setting → Accuracy priority									
	1sec	2sec	5sec	10sec	20sec	30sec	1min	2min	5min	10min
1										
2										
5										
10										
20										
50										
100										
200										
500										
1000										
2000										
5000										
10000										

Example) When measuring the flow rate of 2000 ml/min

Set the calculation time in 1 second to 5 seconds

15. Gross Net Tare Function


Even a small tare in the zero range can be recognized as a tare, you can manage the tare value.

To use this function, enter the Function table as follow, and set "Application Function *AP Fnc*" to "4" in "Application mode *APF*".

15-1 Preparation Of Gross Net Tare Function

Setting gross

- 1 Press and hold the **SAMPLE** key until **bASFnC** of the function table is displayed, then release the key.
- 2 Press the **SAMPLE** key several times to display **AP Fnc**.
- 3 Press the **PRINT** key to display **APF**.
- 4 Press the **RE-ZERO** key several times to display **4 APF GNT**.
- 5 Press the **PRINT** key to store the setting.
- 6 Press the **CAL** key to return to the calculating display.



Key operation

In case of *GNT* setting, operate with the following keys.

ON:OFF key (Long press) Display ON → OFF
 (Short press) Display OFF → ON

RE-ZERO key

- 1 Gross zero (In the "g" display, the range of the minimum scale is "0".)
 Clear a tare value With nothing on the weighing pan, press the **ON:OFF** key and press the **RE-ZERO** key in the state of gross zero (Gross zero mark lit *1)
- 2 When the gross(total amount) is non-zero and positive value.
 Store a tare value(Update)
- 3 When the gross(total amount) is non-zero and negative value.
 Do nothing


Press the **ON:OFF** key

- 1 When it is within the zero range ($\pm 2\%$ from power-on zero)(Refer to *5-1 Basic Operation)
 Update the zero point
- 2 When it is out of the range of zero
 Do nothing

Display

"G", "NET", "PT" mark display


- "G" mark lit When the tare value is zero
- "NET" mark lit When the tare value is out of zero
- "NET" "PT" mark lit When the tare value is set by the PT command (preset tare).




Output

- 1 Every time pressing the PRINT key, it will output in the order of "NET"(object), "GROSS"(total amount), "TARE"(tare).
- 2 By using the UFC function, it can be set the output contents and order.
- 3 Output format is A&D standard format only.


Output example



15-2 Example Of Using The Gross Net Tare Function

- 1 After setting the gross net tare function, press the **ON:OFF** key when nothing is on the weighing pan. "G" will be displayed on the display.
- 2 Place the container to be tared on the weighing pan.
- 3 Press the **RE-ZERO** key to display , the tare value is set (updated). "NET" is displayed on the display.
- 4 Place the object.
- 5 Press the **PRINT** key, it will output in the order of "NET"(object), "GROSS"(total amount), "TARE"(tare).
- 6 Remove anything on the weighing pan and press the **ON:OFF** key to return to the "1".

To continue weighing without changing the tare value, remove the object only, place the next weighing object and press the **PRINT** key to continue outputting.



16. Minimum Weighing Warning Function

The minimum weighing value is the minimum necessary amount of sample to be used for correctly performing quantitatively performing quantitative analysis, taking into consideration measurement error of the balance.

If the amount of sample is too small, the proportion of the measurement error in the measured value increases accordingly, and the reliability of the analysis result may drop.

By using the minimum weighing warning function, it is possible to judge at a glance whether the amount of sample meets the set minimum weight value. This function can be used only in "mg or g" mode.

"*M,N*" is displayed at the top of the unit part when in use.

When the amount of sample is less than the set minimum weighing value, the "*M,N*" indication flashes.

When the amount of sample reaches the minimum weighing value or more, the "*M,N*" indicator will turn off

The minimum weighing value can be changed from the function setting. The factory setting is 0 g.

If the set value is 0 g, no warning will be displayed even if the minimum weighing warning function is ON. Also, a value greater than weighing capacity cannot be set as minimum weighing value.

There are two kinds of warning display as follows

"Excluding near zero"


"Including near zero"

Near zero is within ± 10 digits of 0 g.

Setting procedure

1. Hold down the **SAMPLE** key to display the function setting **bASFnC**.
2. Press the **SAMPLE** key several times to display **MWFnC**.
3. Press the **PRINT** key.
4. **MW-CP** will be displayed. Press the **RE-ZERO** key to change the display from **MW-CP OFF** to **MW-CP Ex0** (excluding near zero) or **MW-CP IN0** (including near zero).
5. To change the setting of the minimum weighing value, proceed to 6. If the minimum weighing value will not be displayed, press the **PRINT** key.
6. Press the **SAMPLE** key to display **MW**.
7. Press the **PRINT** key.
8. Set the minimum weighing value. The minimum weighing value can be changed by the following key operation.


RE-ZERO (+) key	Change the value of the blinking digit.
MODE (-) key	Change the value of the blinking digit.
SAMPLE key	Move the blinking digit.
PRINT key	Store the set value and proceed to the next item.
CAL key	Advance to the next item without storing the setting value.
9. Press the **PRINT** key to return to the weighing display.
10. Press the **CAL** key to return to the weighing display.



Setting confirmation and changing method from the weighing display

1. Press the **MODE** key in the weighing display.
2. The current setting minimum weighing value blinks.
3. Press the **PRINT** key.
4. Set the minimum weighing value. The minimum weighing value can be changed by the following key operation.

RE-ZERO (+) key	Change the value of the blinking digit.
MODE (-) key	Change the value of the blinking digit.
SAMPLE key	Move the blinking digit.
PRINT key	Store the set value and proceed to the next item.
CAL key	Advance to the next item without storing the setting value.



Note

- If *MW-CP* is set to anything other than $\bar{0}$, the unit is fixed in "g" units and it cannot be changed the unit with the **MODE** key.
- To turn OFF the minimum weighing value warning function, set *MW-CP* to $\bar{0}$ by referring to "Setting procedure" above.


17. Underhook

The underhook can be used for magnetic materials or density measurement. The built-in underhook is revealed by removing the plastic cap on the bottom of the balance.

Use the underhook as shown below.

Caution

- Do not apply excessive force to the underhook.
- When not in use, do not open the cover to prevent dust from getting into the balance.
- Do not push the underhook upward.



- The weighing pan, pan support and draft gate fall off, when turning over the balance. Remove them first.
- When not in use, attach the plastic cap to prevent dust from getting into the balance.

18. Programmable-Unit

This is a programmable unit conversion function. It multiplies the weighing data in grams by an arbitrary coefficient set in the function table and displays the result.

The coefficient must be within the range between the minimum and maximum shown below. If the coefficient set is beyond the range, an error is displayed and the balance returns to the coefficient setting mode, prompting to enter an appropriate value. A coefficient of 1 was set at the factory.

Model	Minimum coefficient	Maximum coefficient
GX/GF-203A/303A/403A/603A/1003A/1603A	0.000001	1000
GX/GF-2002A/3002A/4002A/6002A/10002A		100
GX/GF-6001A/10001A		10

Operation

- 1 Press and hold the **SAMPLE** key until **bASFnC** of the function table is displayed.
- 2 Press the **SAMPLE** key several times to display **MLt**.
- 3 Press the **PRINT** key. The balance enters the mode to confirm or set the coefficient.


Confirming the coefficient

- 4 The current coefficient is displayed with the first digit blinking.
 - When it is not to be changed, press the **CAL** key and proceed to step 6.
 - When it is to be changed, press the **RE-ZERO** key and proceed to step 5.

Setting the coefficient

- 5 Set the coefficient using the following keys.

- SAMPLE** key To select a digit to change the value. The selected digit blinks.
- RE-ZERO** key To change the value.
- MODE** key To change the decimal point position. Each time the switch is pressed, the decimal point position changes as follows:



- PRINT** key To store the new setting, display **End** and go to step 6.
- CAL** key To cancel the new setting and go to step 6.

Quitting the operation

- 6 The balance displays **Unit**. Press the **CAL** key to exit the programmable-unit function and return to the weighing mode.

Using the function

Press the **MODE** key to select the programmable-unit (no display on the unit section). Perform weighing as described in “5-1 Basic Operation (Gram Mode)”. After weighing, the balance displays the result (weighing data in grams x coefficient).

19. Density Measurement

The balance is equipped with a density mode. It calculates the density of a solid using the mass value of a sample in air and the mass value in liquid.

For measurement, it use of the option GXA-13 specific gravity measurement kit is remommended.

Note

- The density mode was not selected for use when the balance was shipped from the factory. To use the mode, change the function table and activate the density mode "d5". Refer to "4.Weighing Units".
- Minimum display is fixed while density mode.

Formula to obtain the density

1. Density of solid

It can be obtained from the weight of the sample in air, the weight in the liquid, and the density of the liquid.

$$\rho = \frac{A}{A-B} \times \rho_0$$

ρ : Density of a sample A: Mass value of a sample in air
 ρ_0 : Density of a liquid B: Mass value of a sample in liquid

2. Density of liquid

Weight in air, weight in liquid and volume of float can be obtained using a float of a known.

$$\rho = \frac{A-B}{V}$$

ρ : Density of a sample A: Mass value of a sample in air
 V : Volume of float B: Mass value of a sample in liquid

(1) Prior to measurement: Changing the function table

Prior to measurement, change the function table as follows:

1. Register the density mode.

Density mode cannot be used at the factory setting.

Please refer to "10.Function Table, Unit registration" and register the gravimeter mode (d5).

Density mode is selected as one of the units with the **MODE** key.

2. Select whether the object to be measured is solid or liquid. (Function setting d5 Fnc d5)

3. In the case of solid density measurement, select a method of inputting the density of liquid (function setting d5 Fnc , L d in)

Density of liquid can be set by water temperature input or direct input of density, or input by the following function setting can be selected.

4. To start the measurement, display the weighing display.

Press the **MODE** key to display the specific gravity measurement display.(Refer to the following page "Setting procedure,2,3")

Note

- The following density function (d5 Fnc) is not displayed in the function settings unless density mode is enabled. First, perform the "Register the density mode" operation with the unit setting (L n t) of the function setting. When density mode is activated, "d5 Fnc" appears next to "L n t". For how to change the function setting, refer to "10. Function table".

Class	Item and parameter	Description
d5 Fnc Density function	L d in	Input water temperature
	Liquid density input	Input density directly
	d5	Density measurement of solid
	Measurement object select	Density measurement of liquid

▪ Factory setting

(2) Method of measuring density (specific gravity) of solid (function setting $d5 0$)

Note

- Re-set the density of the liquid with "(3) Entering the density of a liquid" as necessary, such as when the temperature of the liquid changes during measurement or when changing the type of liquid.
- In the density display, the four digits after the decimal point are fixed. The minimum display cannot be changed by pressing the **[SAMPLE]** key.
Density measurement displays the density fixed by measuring the weight in air and measuring the weight in liquid.

The relationship between each state and display is as follows.


Setting procedure

1. Check the weight measurement mode in air (g lights, blink ◀).
Press the **[RE-ZERO]** key to display zero without placing anything on the weighing pan.
2. Place the sample on the weighing pan in air and wait for the display to stabilize. If outputting the mass of the sample, press the **[PRINT]** key.
Next, press the **[SAMPLE]** key to fix the weight in air, and move to the weight measurement mode in liquid (g lights, blink ◀).

- When minus or E is displayed (when weighing capacity is exceeded), press the **[SAMPLE]** key.
- 3. Transfer the sample from the weighing pan in air to the weighing pan in liquid and wait for the display to stabilize. If outputting the mass of the sample, press the **[PRINT]** key.
Next, press the **[SAMPLE]** key to fix the weight in liquid and shift to the density input mode (g turned off, ◀).

Note


- When E is displayed (when weighing capacity is exceeded) press the **[SAMPLE]** key is invalid.
- 5. Enter the density of the liquid.
Refer to "(3) Entering the density of a liquid" and set the density.
Next, press the **[PRINT]** key to enter the density mode (g turns off, ◀).
- 6. If outputting the density, press the **[PRINT]** key.
If measuring another sample, press the **[SAMPLE]** key and start with the weighing mode in air.
The density unit is " $d5$ ".
- 7. Re-set the density of the liquid with "(3) Entering the density of a liquid" as necessary, such as when the temperature of the liquid changes during measurement or when changing the type of liquid.
- 8. Press the **[MODE]** key to enter another weighing mode.



(3) Entering the density of a liquid

Two ways to set the density of a liquid are available in the function table, "Liquid density input (L d in): by entering the water temperature or by entering the density directly.

Entering the water temperature (L d in T)



The water temperature currently set (unit: °C, factory setting : 25°C) is displayed.

Use the following keys to change the value. Setting range is 0.0°C to 99.9°C, in increments of 0.1°C. Refer to the following matrix the " The relation between the water temperature and density

RE-ZERO (+) key.....The key to increase the temperature by one degree.
(0°C is displayed after 99°C)

MODE (-) key.....The key to decrease the temperature by one degree.
(99°C is displayed after 0°C)

SAMPLE key.....Move the blinking digit.

PRINT key.....The key to store new water temperature and return to the density mode.(Proceed to Step 5)

CAL key.....The key to cancel the change and return to the density mode.
Proceed to Step 5.

The relation between the water temperature and density

°C	+0	+1	+2	+3	+4	+5	+6	+7	+8	+9
0	0.99984	0.99990	0.99994	0.99996	0.99997	0.99996	0.99994	0.99990	0.99985	0.99978
10	0.99970	0.99961	0.99949	0.99938	0.99924	0.99910	0.99894	0.99877	0.99860	0.99841
20	0.99820	0.99799	0.99777	0.99754	0.99730	0.99704	0.99678	0.99651	0.99623	0.99594
30	0.99565	0.99534	0.99503	0.99470	0.99437	0.99403	0.99368	0.99333	0.99297	0.99259
40	0.99222	0.99183	0.99144	0.99104	0.99063	0.99021	0.98979	0.98936	0.98893	0.98849

g/cm³

Entering the density directly (L d in) .

The density currently set (unit : g / cm³, factory setting : 1.0000g / cm³) is displayed.



Use the following keys to change the value.

The range to set the density is 0.0000g / cm³ to 1.9999g / cm³.

If it is input beyond the settable range value, **Error 2** is displayed and the display return to the input display.

RE-ZERO (+) key.....The key to set the value of the blinking digit .

MODE (-) key.....The key to select the blinking digit to change the value.

SAMPLE key.....Move the blinking digit.

PRINT key.....The key to store the change and return to the density mode. Proceed to Step 5.

CAL key.....The key to cancel the change and return to the density mode. Proceed to Step 5.

(4) Measuring the density of a liquid (Function table d_5 1)

Density display is four decimal places. Minimum display can not change with the **[SAMPLE]** key.

Density is displayed after "Mass measurement in air" and "Mass measurement in liquid"

The procedure of each measurements is as follows:

Measuring procedure

1. Enter the density mode that "g (gram)" is displayed and the processing indicator (◀) blinks. Place nothing on both pan and press the **[RE-ZERO]** key to display zero.
2. Place the sample on the pan in air.
If the weight value is stored or output, press the **[PRINT]** key to store it after a stable weight value is displayed.
Press the **[SAMPLE]** key to decide the weight value in air and proceed to next step. (g lights, ▶ blinks)


Note

- If negative value or E (out of range) is displayed, the **[SAMPLE]** key is inactive.
Place the liquid to measure the density of in the beaker and sink the float.
At this time, adjust so that the float is about 10 mm below the liquid level.

3. Move the sample to the pan in liquid.
If the weight value is stored or output, press the **[PRINT]** key to store it after a stable weight value is displayed. Press the **[SAMPLE]** key to decide the weight value in liquid and proceed to next step. (g turned off, $\frac{1}{10d}$ lights, ▶ lights)

Note

- If E (out of range) is displayed, the **[SAMPLE]** key is inactive.
4. Enter the volume of the float.
Refer to "(5) Entering the volume of the float" and enter.
Then press the **[PRINT]** key to return to the density mode.
 5. If the density value is stored or output, press the **[PRINT]** key to store it. If the other sample is measured, press the **[SAMPLE]** key, and start from measurement of weighing mode in the air. The density unit is " $\frac{1}{15}$ ".
6. Press the **[MODE]** key to proceed to other modes.




(5) Entering the volume of the float

The volume of the float that is currently set is displayed. (Factory setting is 10.00 c m³)

Change the setting value as follows.

The setting range is 0.01 c m³ to 99.99 c m³, every 0.01 c m³.



RE-ZERO (+) Key... The key to set the value of the blinking digit .

MODE (–) key..... The key to select the blinking digit to change the value.

SAMPLE key..... Move the blinking digit.

PRINT key..... The key to store the change and return to the density mode. Proceed to Step 5.

CAL key..... The key to cancel the change and return to the density mode. Proceed to Step 5.

20.Password Lock Function

The password function can be specified for an authorized user by using a password stored in the balance by manager. While the password function is used, the balance can not be weighed without inputting the correct password.

There are two login levels : administrator (AdMin.) and user (oPR.1 ~ 10).

User can individually set the password for 10 people. Administrators can use all functions and settings.

User can only perform weighing.

Users cannot change function settings including initialization and clock setting.

Login level	Weighing					Select function and initialization	Function setting (Date/time)
	Weighing value Display · Output	Re-zero	Calibration	Change the unit	Change the minimum display		
Administrator (AdMin.)	Possible	Possible	Possible	Possible	Possible	Possible	Possible
Operator (oPR.1~10)	Possible	Possible	Impossible	Impossible	Impossible	Impossible	Impossible

Password: Four digits (4 x 4 x 4 x 4 = 256 outcomes) using [MODE], [SAMPLE], [PRINT] and [RE-ZERO] keys. At factory setting, the password function is not used.

The management of the password function is performed in "Password function" of the Function table.


20-1 Using Password Function

The password function can be switched between "Use" or "Not used" at "Password function (PASS Wd)" of the Function Table

- Press and hold the [SAMPLE] key in the weighing mode.
[bR5Fnc] is displayed.
- Press the [SAMPLE] key several times until [PASSWd] is displayed.
- Press the [PRINT] key to display [LOCK OFF].
To cancel the operation, press the [CAL] key.
- Press the [RE-ZERO] key to display [LOCK ON].
- Press the [PRINT] key to display [SurE :YES No] where "No" is selected.
(where "No" is blinking when "No" is selected.)
- Press the [RE-ZERO] to change YES / No .
- Display the [SurE :YES No] .
(where "Yes" is blinking when "Yes" is selected.)
- Press the [PRINT] key when "Yes" is selected, the password function can now be used.

When turning on the balance, input the new password.

- [PASSNo.] is displayed. To register or change the password, proceed the next page "2". If it is not to register, press the [CAL] key twice to return to the weighing display.



20-2 Changing Password

The password can be changed at "Password (*PASSNo.*)" of the Function Table.

1. Press and hold the **[SAMPLE]** key in the weighing mode.
bASFnC is displayed.
2. Press the **[SAMPLE]** key several times until PASSWd is displayed.
3. Press the **[PRINT]** key to display LockON. If the display is OFF, press **[RE-ZERO]** to switch to ON and press the **[PRINT]** to set the displayed setting.
4. PASSNo. is displayed.
5. Press the **[PRINT]** key to display the login level (*AdMin.*).
6. Press the **[SAMPLE]** key to change the login level
AdMin. / oPR. 1~10.

If the password is already registered at the login level, the stability mark ° is lit. (changeable)

7. Press the **[PRINT]** key to display the current password.
 At factory settings, the password is 7777.
 (the **[RE-ZERO]** key, four times)
8. Set the new password using the following keys.


The balance will turn automatically after no operation for ten minutes.

[MODE] key	Character	<i>M</i>
[SAMPLE] key	Character	<i>S</i>
[PRINT] key	Character	<i>P</i>
[RE-ZERO] key	Character	<i>Z</i>
[CAL] key	Back key	
[CAL] key(long press)	..	Delete password	

9. Input four characters of the new password using these keys.
10. The balance displays SurE :yESNo where "No" is blinking when "No" is selected.
 (Press the **[CAL]** key to return to the 4th character input.)
11. Press the **[RE-ZERO]** key to display SurE :yESNo where "YES" is blinking when "YES" is selected.
12. Press the **[PRINT]** key to store the new password when displaying SurE :yESNo.
 When turning on the balance, input the new password.


Note

- If the correct password is missing, the balance can not be used. Record the correct password after changing.
- Passwords that have already been registered by administrator (*AdMin.*) can not be registered by users (*oPR. 1~10*).



How to delete the password (OPR. 1~10)

1. Press and hold the **CAL** key to display and blink the **CLEAR** when setting the password.
2. Press the **PRINT** key to display **CLEAR No.**
3. Press the **RE-ZERO** to change **Go/No.**
4. Press the **PRINT** when **CLEAR Go** to delete the password.



20-3 Inputting Password When Turning On The Balance


When turning on the balance with the password function ON, the password input is displayed before the weighing display.

1. Turn on the balance using the **ON:OFF** key.
2. **PASSword** is displayed and **.....PASS** of the password input is displayed.

3. Input the four characters of the correct password using the following keys.

MODE key.....	Character	M
SAMPLE key.....	Character	S
PRINT key.....	Character	P
RE-ZERO key.....	Character	Z
CAL key.....	Back key	

4. When the password is correct, the balance displays weighing value after full-segment display.
When the password is incorrect, **FAIL** is displayed and the buzzer sounds three times.



20-4 Missing Password

If the correct password is missing, the balance can not be used.
Contact your local A&D dealer to reset the password to factory settings.

21.Maintenance

21-1 Treatment Of The Balance

- Clean the balance with a lint free cloth that is moistened with warm water and a mild detergent.
- Do not use organic solvents to clean the balance.
- Do not disassemble the balance.
- Use the original packing material for transportation.

22.Troubleshooting

22-1 Checking The Balance Performance And Environment

The balance is a precision instrument. When the operating environment or the operating method is inadequate, correct weighing can not be performed. Place a sample on the pan and remove it, and repeat this several times. If the balance seems to have a problem with repeatability or to perform improperly, check as described below. If improper performance persists after checking, contact the local A&D dealer for repair.

"Frequently asked questions" and their answers are also posted on our website

<<http://www.aandd.co.jp>>.

1. Checking that the balance performs properly

- Please check the operation of the balance by the self diagnosis function. Refer to "6-2 Self-diagnosis Function". Fatal faults are indicated by messages.
- Check the balance performance using an external weight. Be sure to place the weight in the center of the weighing pan.
- Check the balance repeatability, linearity and calibrated value using external weights with a known value.

2. Checking that the operating environment or weighing method is proper

Operating environment

- Is the weighing table solid enough? (Especially 1mg model)
- Is the balance level? Refer to " 2-2 Precaution "How to adjust the bubble spirit level".
- Is the operating environment free from vibration and drafts?
- Is there a strong electrical or magnetic noise source such as a motor near the balance?












Weighing method















- Does the weighing pan rim touch anything? Is the weighing pan assembly installed correctly?
- Is the **RE-ZERO** key pressed before placing a sample on the weighing pan?
 - Is the sample placed in the center of the weighing pan?
 - Has the balance been calibrated using the internal mass (one-touch calibration)? (Only for GX=A series)
- Has the balance been warmed up for one hour before weighing?

Sample and container

- Has the sample absorbed or lost moisture due to the ambient conditions such as temperature and humidity?
- Has the temperature of the container been allowed to equalize to the ambient temperature? Refer to "2-3 During Use".
- Is the sample charged with static electricity? Refer to "2-3 During Use".
(This occurs especially with 0.001g models when the relative humidity is low.)
- Is the sample of magnetic material such as iron? There are cautions about weighing magnetic materials. Refer to "2-3 During Use".

22-2 Error Codes

Display	Error code	Description
		<p>Data abnormality of the internal sensor of the balance</p> <ul style="list-style-type: none"> <input type="checkbox"/> If this error occurs during weighing and you return to the normal weighing state after a certain period of time, there is a possibility that the data temporarily became abnormal due to static electric noise if you are measuring an easily chargeable item. ※ If the sample to be measured is charged, it is recommended that you use a AD- 1683 static eliminator to discharge the sample before measurement in order to more accurately weigh the sample. <input type="checkbox"/> If this error continues to be displayed and can not be resolved, the internal sensor or the circuit may be damaged. Please contact your dealer.
	E C, E 1 1	<p>Stability error</p> <p>The balance cannot stabilize due to an environmental problem. Check around the pan. Prevent vibration, drafts, temperature changes, static electricity and magnetic fields, from influencing the balance. Refer to "2-3 During Use".</p> <p>To return to the weighing mode, press the  key.</p>
		<p>Out of the setting range</p> <p>The data to be stored is out of the setting range.</p>
		<p>Malfunction of the internal memory element of the balance</p> <p>If this error continues to be displayed, repair is necessary. Please contact your dealer.</p>
	E C, E 1 6	<p>Internal mass error</p> <p>Applying the internal mass does not yield a change in the mass value as specified.</p> <p>Confirm that there is nothing on the pan and perform the weighing operation from the beginning again.</p>
	E C, E 1 7	<p>Internal mass error</p> <p>The internal mass application mechanism does not function properly. Perform the weighing operation from the beginning again.</p>
		<p>Abnormality in the internal memory data of the balance</p> <p>If this error continues to be displayed, repair is necessary. Please contact your dealer.</p>
		<p>Abnormality in the internal memory data of the balance</p> <p>If this error continues to be displayed, repair is necessary. Please contact your dealer.</p>
	E C, E 2 0	<p>Calibration weight error</p> <p>The calibration weight is too heavy. Confirm the calibration mass value. Press the  key to return to the weighing</p>

Display	Error code	Description
		mode.
	E C, E 2 1	Calibration weight error The calibration weight is too light. Confirm the calibration mass value. Press the  key to return to the weighing mode.
		Overload error A sample beyond the balance weighing capacity has been placed on the pan. Remove the sample from the pan.
		Weighing pan Error The mass value is too light. Confirm that the weighing pan is properly installed and calibrate the balance.
		Sample mass error The balance can not store the sample for the counting mode or for the percent mode because it is too light. Use a larger sample.
  		Unit mass error The sample unit mass for the counting mode is too light. Storing and using it for counting will cause a counting error. Add samples to reach the specified number and press the  key. Pressing the  key without adding samples will shift the balance to the counting mode. But, for accurate counting, be sure to add samples.
		Clock battery error The clock backup battery has been depleted. Press any key and set the time and date. The clock and calendar function works normally as long as the AC adapter is connected to the balance. If this error appears frequently, contact the local A&D dealer.
		Power supply voltage fault The voltage supplied from the AC adapter is abnormal. Please check if the problem is the AC adapter (TB248) supplied with the balance.
		Full memory The maximum number of stored weighing values has been reached. In order to store more weighing values, it is necessary to delete the data. Refer to "11.Data Memory"
		Full memory The stored calibration history has reached 50 instances. If more is stored, the old history will be deleted. Refer to "11.Data Memory"
	E C, E 0 0	Communications error A protocol error occurred in communications. Confirm the format, baud rate and parity.
	E C, E 0 1	Undefined command error

Display	Error code	Description
		An undefined command was received. Confirm the command.
	E C, E 0 2	Not ready A received command can not be processed. Example: <ul style="list-style-type: none"> □ The balance received a "Q" command, but not in the weighing mode. □ The balance received a "Q" command while processing a RE-ZERO command. Adjust the delay time to transmit a command.
	E C, E 0 3	Timeout error If the timeout parameter is set to " <i>t-UP</i> ", the balance did not receive the next character of a command within the time limit of one second. Confirm the communication.
	E C, E 0 4	Excess characters error The balance received excessive characters in a command. Confirm the command.
	E C, E 0 6	Format error A command includes incorrect data. Example: <ul style="list-style-type: none"> □ The data is numerically incorrect. Confirm the command.
	E C, E 0 7	Parameter setting error The received data exceeds the range that the balance can accept. Confirm the parameter range of the command.

22-3 Other Display



When this indicator (◀) blinks, automatic self calibration is required. The indicator blinks when the balance detects changes in ambient temperature. If the balance is not used for several minutes with this indicator blinking, the balance performs automatic self calibration. The blinking period depends on the operating environment.

Advise The balance can be used while this indicator is blinking. We recommend that you perform automatic self calibration for precision weighing.

22-4 Asking For Repair

If the balance needs service or repair, contact your local A&D dealer.

The balance is a precision instrument. Use much care when handling the balance and observe the following when transporting the balance.

- Use the original packing material for transportation.
- Remove the weighing pan, pan support, breeze break ring and dust plate from the main unit.

23. Specifications

23-1 GX-A series 0.001g models

		GX-1603A	GX-1003A	GX-603A	GX-403A	GX-303A	GX-203A
Weighing capacity		1620g	1100g	620g	420g	320g	220g
Maximum display		1620.084g	1100.084g	620.084g	420.084g	320.084g	220.084g
Minimum display		0.001g					
Repeatability (Standard deviation)		0.002g(1600g) 0.001g(1000g)	0.001g				
Linearity		±0.003g			±0.002g		
Stabilization time (FAST setting, good environment)		1600g:Approx. 1.5 sec 5g:Approx.0.8sec	Approx.1sec (50g: Approx. 0.8 sec)				
Sensitivity drift		±2ppm/°C (Automatic self calibration OFF)					
Accuracy after calibration with internal mass※		±0.010g (1000g)	±0.010g				
Operating environment		5°C to 40°C, 85%RH or less (No condensation)					
Display refresh rate		5 times/second or, 10 times/second or 20 times/second					
Counting mode	Minimum unit mass	0.001g					
	Number of samples	5,10,25,50 or 100 pieces					
Percent mode	Minimum 100% mass	0.100g					
	Minimum 100% display	0.01 %, 0.1 %, 1 % (Depends on the reference mass stored.)					
Carat	Weighing capacity	8100ct	5500ct	3100ct	2100ct	1600ct	1100ct
	Minimum display	0.05ct					
Monme	Weighing capacity	432mom	293mom	165mom	112mom	85mom	58mom
	Minimum display	0.005mom					
Interface		RS-232C, USB					
External calibration weight		50g 100g (100g interval) 1600g	50g 100g (100g interval) 1000g	50g 100g (100g interval) 600g	50g 100g (100g interval) 400g	50g 100g 200g 300g	50g 100g 200g
Weighing pan		128mm x 128mm					
External dimensions		212 (w) x317 (D) x93 (H)					
Power supply (AC adapter)		Power consumption: Approx. 30VA (supplied to the AC adapter) Confirm that the adapter type is correct for the local voltage and power receptacle type.					
Weight		Approx.5kg					

※ The operating environment does not include excessive change of ambient temperature, humidity, vibration, drafts, magnetic fields and static electricity.

The internal mass may change due to corrosion or other damage caused by the operating environment, or due to aging.

23-2 GX-A series 0.01g models

		GX-10002A	GX-6002A	GX-4002A	GX-3002A	GX-2002A
Weighing capacity		10200g	6200g	4200g	3200g	2200g
Maximum display		10200.84g	6200.84g	4200.84g	3200.84g	2200.84g
Minimum display		0.01g				
Repeatability (Standard deviation)		0.02g(10000g) 0.01g(5000g)	0.01g			
Linearity		±0.03g		±0.02g		
Stabilization time (FAST setting, good environment)		10kg: Approx. 1.5 sec 50kg: Approx. 0.8 sec	Approx.1 sec (50g: Approx. 0.8 sec)			
Sensitivity drift		±2ppm/°C (Automatic self calibration:OFF)				
Accuracy after calibration with internal mass※		±0.15g (5000g)		±0.15g		±0.10g
Operating environment		5°C to 40°C, 85%RH or less (No condensation)				
Display refresh rate		5 times/second or, 10 times/second or 20 times/second				
Counting mode	Minimum unit mass	0.01g				
	Number of samples	5,10,25,50 or 100 pieces				
Percent mode	Minimum 100% mass	1.00g				
	Minimum 100% display	0.01 %, 0.1 %, 1 % (Depends on the reference mass stored.)				
Carat	Weighing capacity	50100ct	31000ct	21000ct	16000ct	11000ct
	Minimum display	0.05ct				
Monme	Weighing capacity	2720mom	1653mom	1120mom	853mom	586mom
	Minimum display	0.05mom				
Interface		RS-232C、USB				
External calibration weight		500g 1000g (1000g interval) 10000g	500g 1000g (1000g interval) 6000g	500g 1000g (1000g interval) 4000g	500g 1000g 2000g 3000g	500g 1000g 2000g
Weighing pan		165mm x 165mm				
External dimensions		212 (w) x317 (D) x93 (H)				
Power supply (AC adapter)		Power consumption: Approx. 30VA (supplied to the AC adapter) Confirm that the adapter type is correct for the local voltage and power receptacle type.				
Weight		5kg				

※ The operating environment does not include excessive change of ambient temperature, humidity, vibration, drafts, magnetic fields and static electricity.

The internal mass may change due to corrosion or other damage caused by the operating environment, or due to aging.

23-3 GX-A series 0.1g models

		GX-10001A	GX-6001A
Weighing capacity		10200g	6200g
Maximum display		10208.4g	6208.4g
Minimum display		0.1g	
Repeatability (Standard deviation)		0.1g	
Linearity		±0.1g	
Stabilization time (FAST setting, good environment)		Approx. 1 seconds (500g: Approx.0.8seconds)	
Sensitivity drift		±2ppm/°C (Automatic self calibration:OFF)	
Accuracy after calibration with internal mass※		±0.5g (5000g)	
Operating environment		5°C to 40°C, 85%RH or less (No condensation)	
Display refresh rate		5 times/second or, 10 times/second or 20 times/second	
Counting mode	Minimum unit mass	0.1g	
	Number of samples	5,10,25,50 or 100 pieces	
Percent mode	Minimum 100% mass	10.0g	
	Minimum 100% display	0.01 %, 0.1 %, 1 % (Depends on the reference mass stored.)	
Carat	Weighing capacity	50100ct	31000ct
	Minimum display	0.5ct	
Monme	Weighing capacity	2720mom	1653mom
	Minimum display	0.5mom	
Interface		RS-232C, USB	
External calibration weight		500g 1000g (1000g interval) 10000g	500g 1000g (1000g interval) 6000g
Weighing pan		165mm x 165mm	
External dimensions		212 (w) x317 (D) x93 (H)	
Power supply (AC adapter)		Power consumption: Approx. 30VA (supplied to the AC adapter) Confirm that the adapter type is correct for the local voltage and power receptacle type.	
Weight		5kg	

※ The operating environment does not include excessive change of ambient temperature, humidity, vibration, drafts, magnetic fields and static electricity.

The internal mass may change due to corrosion or other damage caused by the operating environment, or due to aging.

23-4 GF-A series 0.001g models

	GF-1603A	GF-1003A	GF-603A	GF-403A	GF-303A	GF-203A	
Weighing capacity	1620g	1100g	620g	420g	320g	220g	
Maximum display	1620.084g	1100.084g	620.084g	420.084g	320.084g	220.084g	
Minimum display	0.001g						
Repeatability (Standard deviation)	0.002g(1600g) 0.001g(1000g)	0.001g					
Linearity	±0.003g		±0.002g				
Stabilization time (FAST setting, good environment)	1600g: Approx. 1.5 sec 5g: Approx. 0.8 sec	Approx.1 sec (5g: Approx.0.8sec)					
Sensitivity drift	±2ppm/°C						
Operating environment	5°C to 40°C, 85%RH or less (No condensation)						
Display refresh rate	5 times/second or, 10 times/second or 20 times/second						
Counting mode	Minimum unit mass	0.001g					
	Number of samples	5,10,25,50 or 100 pieces					
Percent mode	Minimum 100% mass	0.100g					
	Minimum 100% display	0.01 %, 0.1 %, 1 % (Depends on the reference mass stored.)					
Carat	Weighing capacity	8100ct	5500ct	3100ct	2100ct	1600ct	
	Minimum display	0.05ct					
Monme	Weighing capacity	432mom	293mom	165mom	112mom	85mom	58mom
	Minimum display	0.005mom					
Interface	RS-232C, USB						
External calibration weight	50g 100g (100g interval) 1600g	50g 100g (100g interval) 1000g	50g 100g (100g interval) 600g	50g 100g (100g interval) 400g	50g 100g 200g 300g	50g 100g 200g	
Weighing pan	128mm x 128mm						
External dimensions	212 (w) x317 (D) x93 (H)						
Power supply (AC adapter)	Power consumption: Approx. 30VA (supplied to the AC adapter) Confirm that the adapter type is correct for the local voltage and power receptacle type.						
Weight	5kg						

23-5 GF-A series 0.01g models


		GF-10002A	GF-6002A	GF-4002A	GF-3002A	GF-2002A
Weighing capacity		10200g	6200g	4200g	3200g	2200g
Maximum display		10200.84g	6200.84g	4200.84g	3200.84g	2200.84g
Minimum display		0.01g				
Repeatability (Standard deviation)		0.02g(10000g) 0.01g(5000g)	0.01g			
Linearity		±0.03g		±0.02g		
Stabilization time (FAST setting, good environment)		10kg: Approx. 1.5 sec 50g: Approx. 0.8 sec	Approx.1 seconds (50g: Approx.0.8seconds)			
Sensitivity drift		±2ppm/°C				
Accuracy after calibration with internal mass※		±0.15g (5000g)	±0.15g		±0.10g	
Operating environment		5°C to 40°C, 85%RH or less (No condensation)				
Display refresh rate		5 times/second or, 10 times/second or 20 times/second				
Counting mode	Minimum unit mass	0.01g				
	Number of samples	5,10,25,50 or 100 pieces				
Percent mode	Minimum 100% mass	1.00g				
	Minimum 100% display	0.01 %, 0.1 %, 1 % (Depends on the reference mass stored.)				
Carat	Weighing capacity	50100ct	31000ct	21000ct	16000ct	11000ct
	Minimum display	0.05ct				
Monme	Weighing capacity	2720mom	1653mom	1120mom	853mom	586mom
	Minimum display	0.05mom				
Interface		RS-232C、USB				
External calibration weight		500g 1000g (1000g interval) 10000g	500g 1000g (1000g interval) 6000g	500g 1000g (1000g interval) 4000g	500g 1000g 2000g 3000g	500g 1000g 2000g
Weighing pan		165mm x 165mm				
External dimensions		212 (w) x317 (D) x93 (H)				
Power supply (AC adapter)		Power consumption: Approx. 30VA (supplied to the AC adapter) Confirm that the adapter type is correct for the local voltage and power receptacle type.				
Weight		5kg				

23-6 GF-A series 0.1g models

	GF-10001A	GF-6001A
Weighing capacity	10200g	6200g
Maximum display	10208.4g	6208.4g
Minimum display	0.1g	
Repeatability (Standard deviation)	0.1g	
Linearity	±0.1g	
Stabilization time (FAST setting, good environment)	Approx. 1 sec (500g: Approx.0.8 sec)	
Sensitivity drift	±2ppm/°C (Automatic self calibration OFF)	
Accuracy after calibration with internal mass※	±0.5g (5000g)	
Operating environment	5°C to 40°C, 85%RH or less (No condensation)	
Display refresh rate	5 times/second or, 10 times/second or 20 times/second	
Counting mode	Minimum unit mass	0.1g
	Number of samples	5,10,25,50 or 100 pieces
Percent mode	Minimum 100% mass	10.0g
	Minimum 100% display	0.01 %, 0.1 %, 1 % (Depends on the reference mass stored.)
Carat	Weighing capacity	50100ct
	Minimum display	0.5ct
Monme	Weighing capacity	2720mom
	Minimum display	0.5mom
Interface	RS-232C、USB	
External calibration weight	500g 1000g (1000g interval) 10000g	500g 1000g (1000g interval) 6000g
Weighing pan	165mm x 165mm	
External dimensions	212 (w) x317 (D) x93 (H)	
Power supply (AC adapter)	Power consumption: Approx. 30VA (supplied to the AC adapter) Confirm that the adapter type is correct for the local voltage and power receptacle type.	
Weight	Approx.5kg	




24. External Dimension

GX-203A / GX-303A / GX-403A / GX-603A / GX-1003A / GX-1603A
 GF-203A / GF-303A / GF-403A / GF-603A / GF-1003A / GF-1603A



- ※1 Opening width when transparent plate is removed.
- ※2 Inside dimension
- ※3 Weighing pan size
- ※4 Height from the weighing pan to the lid of the breeze break.
- ※5 Opening height when transparent plate is removed.
- ※6 Position under the floor weighing platform.
- ※7 DC jack protruding dimension of AC adapter.

GX-2002A / GX-3002A / GX-4002A / GX-6002A / GX-10002A/ GX-6001A/ GX-10001A
 GF-2002A / GF-3002A / GF-4002A / GF-6002A / GF-10002A/ GF-6001A/ GF-10001A



- ※3 Weighing pan size
- ※6 Position under the floor weighing platform.
- ※7 DC jack protruding dimension of AC adapter.

24-1 Options And Peripheral Instruments

Options

GXA-03 : RS-232C interface insulation type

- RS-232C Interface insulation type for expansion

GXA-04 : Comparator output (Relay / with a Buzzer)

- Outputs comparator results.

GXA-06 : Analog voltage output

- This option outputs a voltage of 0 to1V (or 0.2 to1V).

GXA-08 : Ethernet interface

- Enables the balance to communicate with computers on a network.
- Multiple balances on a network can be controlled by one computer.

GXA-09 : Built-in battery unit (Can be installed only at shipment)

- Enables the balance to be used in an environment where the AC adapter can not be used.
- Charging time is approximately 10 hours, and continuous use time is approximately 14 hours.

GXA-10 : Glass breeze break

- Breeze break unit with a glass door

GXA-12 : Animal container kit

- Container with depth to make it difficult for animals to edcape

GXA-13 : Specific gravity measurement kit

- Unit that enables easy weighing of the sample's weight in air and in water.

GXA-23-PRINT : Foot switch input for PRINT

- External contact input terminal that can operate PRINT and RE-ZERO key.
Foot switch of print function included. (AX-SW137-PRINT)

GXA-23-RE-ZERO : Foot switch input for RE-ZERO

- External contact input terminal that can operate PRINT and RE-ZERO key.
Foot switch of RE-ZERO function included. (AX-SW137-REZERO)

GXA-23-PLUG : External input interface

- External contact input terminal that can operate PRINT and RE-ZERO key.
Assembled stereo plug included.

Note: In order to use, it is necessary to solder the attached plug and the switch prepared by the customer.

GXA-24 : USB host interface (Can be installed only at shipment)

- Stores the weighing value in the USB memory.

GXA-25 : Static eliminator for GX-A/GF-A series

- In addition to being used alone, this static eliminator unit can be built in to a large breeze break.

GXA-26 : External IR switch

- External non-contact switch that can operate PRINT and RE-ZERO key.

AX-GXA-31 : Body cover (5 pieces)

- Protective cover for standard accessories

Peripheral devices

AD-8920A : Remote Display

- This option can be connected to the balance using the RS-232C interface or current loop and displays the weighing data transmitted by the balance.

AD-8922A : Remote Controller

- This option can be connected to the balance using the RS-232C interface and can control the balance remotely.

AD-8127 : Compact printer

- Small dot impact printer that connects with the balance via the RS-232C interface.
- Statistical function, clock and calendar function, interval print function, graphic print function, dump print mode

AD-1687 : Weighing Environment logger

- A data logger equipped with 4 sensors for temperature, humidity, barometric pressure and vibration that can measure and store environmental data. When connected to the RS-232C interface of the balance, the AD-1687 can store environmental data along with weighing data. Therefore, it is possible to store data in an environment where a computer can not be used.
- The stored data can be read to a personal computer using USB. As the AD-1687 is recognized as USB memory, special software is not required to read the data.

AD-1688 : Data Logger

- When connected to the RS-232C interface of the balance, the AD-1688 can store the data in an environment where a personal computer can not be used.

AD-1689 : Tweezers for calibration weight

- A pair of tweezers ideally suited for holding calibration weights of 1g to 500g.

AX-USP-9P : USB Converter

- An RS-232C cable is provided to connect the USB converter to the balance.
- Enables bi-directional communication between the PC and the balance when a USB driver is installed.

AX-SW137-PRINT : Foot switch for print (with connector)

- Foot switch that functions in the same way as the PRINT key when combined with GXA-23 external connector

AX-SW137-REZERO : Foot switch for re-zero (with connector)

- Foot switch that functions in the same way as the RE-ZERO key when combined with GXA-23 external connector



A&D Company, Limited

3-23-14 Higashi-Ikebukuro, Toshima-ku, Tokyo 170-0013, JAPAN
Telephone: [81] (3) 5391-6132 Fax: [81] (3) 5391-6148

A&D ENGINEERING, INC.

1756 Automation Parkway, San Jose, California 95131, U.S.A.
Telephone: [1] (408) 263-5333 Fax: [1] (408) 263-0119

A&D INSTRUMENTS LIMITED

Unit 24/26 Blacklands Way, Abingdon Business Park, Abingdon, Oxfordshire OX14 1DY United Kingdom
Telephone: [44] (1235) 550420 Fax: [44] (1235) 550485

A&D AUSTRALASIA PTY LTD

32 Dew Street, Thebarton, South Australia 5031, AUSTRALIA
Telephone: [61] (8) 8301-8100 Fax: [61] (8) 8352-7409

A&D KOREA Limited

한국에이.엔.디(주)
대한민국 서울시 영등포구 여의도동 36-2 맨하탄 빌딩 8층 우편 번호 150-749
(Manhattan Building 8th Floor, 36-2 Yoido-dong, Youngdeungpo-gu, Seoul, 150-749 Korea)
전화: [82] (2) 780-4101 팩스: [82] (2) 782-4280

ООО A&D RUS

ООО "ЭЙ энд ДИ РУС"
121357, Российская Федерация, г.Москва, ул. Верейская, дом 17
(Business-Center "Vereyskaya Plaza-2" 121357, Russian Federation, Moscow, Vereyskaya Street 17)
тел.: [7] (495) 937-33-44 факс: [7] (495) 937-55-66

A&D INSTRUMENTS INDIA PRIVATE LIMITED

ऐ&डी इन्स्ट्रुमेन्ट्स इण्डिया प्रा० लिमिटेड
509, उद्योग विहार, फेस -5, गुडगांव - 122016, हरियाणा, भारत
(509, Udyog Vihar, Phase-V, Gurgaon - 122 016, Haryana, India)
फोन : 91-124-4715555 फैक्स : 91-124-4715599