### Concentrator plus/Vacufuge® plus – Instructions for Setting Up the Device

These Instructions supplement the Concentrator plus/Vacufuge<sup>®</sup> plus operating manual. They do not replace it. Be sure to read the operating manual before you begin to use the device.

#### **Transport**

The device weighs 31 kg (complete system) or 17 kg (basic device).


- Transport the device only when it is in its original packaging.
- ▶ Two persons are always required to transport it.
- ▶ Use a dolly or other mechanical help for longer distances.

#### **Setting Up**

► Follow these steps in order; two persons are required.


#### 1. Open the box.

- a) Cut the straps.
- b) Cut open the tape.
- c) Open all 4 box flaps.
- d) Remove the paperboard covering.
- e) Take out the accessories.



#### 2. Take the device out of the box.

- a) With its transport protection pads still around it, lift the device out of the box.
- b) Carefully lay the device on its side, with transport protection pads.
- c) Gently slide back the rear transport protection pad until the second transport protection screw is visible.




- 1 Rear transport protection pad
- 2 Transport protection screws
- 3 Bottom transport protection pad
- 4 Front transport protection pad

# eppendorf

## Concentrator plus/Vacufuge® plus – Instructions for Setting Up the Device

#### 3. Remove the transport protection pads.

- a) Pull out the bottom piece of transport protection pad.
- b) Unscrew and remove both transport protection screws.
- c) Set the device upright.
- d) Remove the front and rear transport protection pads.
- e) Remove the plastic wrapping from the device.
- f) Remove the plastic wrapping from the cover.
- g) Set the device on a stable, level, and resonance-free laboratory workbench. Let it rest 3 hours to acclimatize.



Retain the box as well as the transport protection pads and screws to transport or store the device later.



#### Caution: It is dangerous to ignore the operating manual.

The device is not yet ready to operate. You must attach further components.

▶ Be sure to read the operating manual included here before you begin to use the device.

# eeppendorf<sup>®</sup> is a registered trademark. 5399 950.017-05/0408 · Printed in Germany

## **eppendorf**

#### In touch with life

Your local distributor: www.eppendorf.com/worldwide

Eppendorf AG · 22331 Hamburg · Germany · Tel: +49 40 538 01-0 · Fax: +49 40 538 01-556 E-Mail: eppendorf@eppendorf.com

Eppendorf North America, Inc. · One Cantiague Road · P.O. Box 1019 · Westbury, N.Y. 11590-0207 · USA Tel: +1 516 334 7500 · Toll free phone: +1 800 645 3050 · Fax: +1 516 334 7506 · E-Mail: info@eppendorf.com

#### **Application Support**

Europe, International: Tel: +49 1803 666 789 · E-Mail: support@eppendorf.com North America: Tel: +1 800 645 3050 ext. 2258 · E-Mail: support\_na@eppendorf.com Asia, Pacific: Tel: +60 3 8023 6869 · E-Mail: support\_asiapacific@eppendorf.com