



# Gilson Pipette Specifications

Systematic Error = Accuracy

Random Error = Precision

## PIPETMAN Classic™

		Pipette Specifications			
Model Part Number Volume Range	Volume (µL)	Systematic Error (µL)	Random Error (µL)	Systematic Error %	Random Error %
P2 F144801 0.2-2 µL	Min. 0.2	± 0.024	≤ 0.012	12%	6%
	0.5	± 0.025	≤ 0.012	5%	2.4%
	Max. 2	± 0.030	≤ 0.014	1.5%	0.7%
P10 F144802 1-10µL	Min. 1	± 0.025	≤ 0.012	2.5%	1.2%
	5	± 0.075	≤ 0.030	1.5%	0.6%
	Max. 10	± 0.100	≤ 0.040	1%	0.4%
P20 F123600 2-20µL	Min. 2	± 0.10	≤ 0.03	5%	1.5%
	5	± 0.10	≤ 0.04	2%	0.8%
	10	± 0.10	≤ 0.05	1%	0.5%
	Max. 20	± 0.20	≤ 0.06	1%	0.3%
P100 F123615 20-100µL	Min. 20	± 0.35	≤ 0.10	1.75%	0.5%
	50	± 0.40	≤ 0.12	0.8%	0.24%
	Max. 100	± 0.80	≤ 0.15	0.8%	0.15%
P200 F123601 50-200µL	Min. 50	± 0.50	≤ 0.20	1%	0.4%
	100	± 0.80	≤ 0.25	0.8%	0.25%
	Max. 200	± 1.60	≤ 0.30	0.8%	0.15%
P1000 F123602 200-1000µL	Min. 200	± 3	≤ 0.6	1.5%	0.3%
	500	± 4	≤ 1.0	0.8%	0.2%
	Max. 1000	± 8	≤ 1.5	0.8%	0.15%
P5000 F123603 1000-5000µL	Min. 1000	± 12	≤ 3	1.2%	0.3%
	2000	± 12	≤ 5	0.6%	0.25%
	Max. 5000	± 30	≤ 8	0.6%	0.16%
P10 mL F161201 1-10mL	Min. 1 mL	± 30	≤ 6	3%	0.6%
	2 mL	± 30	≤ 6	1.5%	0.3%
	5 mL	± 40	≤ 10	0.8%	0.2%
	Max. 10 mL	± 60	≤ 16	0.6%	0.16%

## PIPETMAN G

		Pipette Specifications			
Model Part Number Volume Range	Volume (µL)	Systematic Error (µL)	Random Error (µL)	Systematic Error %	Random Error %
P2G F144054M/P 0.2-2 µL	Min. 0.2	± 0.024	≤ 0.012	12%	6%
	0.5	± 0.025	≤ 0.012	5.0%	2.4%
	Max. 2	± 0.030	≤ 0.014	1.5%	0.7%
P10G F144055M/P 1-10µL	Min. 1	± 0.025	≤ 0.012	2.5%	1.2%
	5	± 0.075	≤ 0.030	1.5%	0.6%
	Max. 10	± 0.100	≤ 0.040	1%	0.4%
P20G F144056M/P 2-20µL	Min. 2	± 0.10	≤ 0.03	5%	1.5%
	10	± 0.10	≤ 0.05	1%	0.5%
	Max. 20	± 0.20	≤ 0.06	1%	0.3%
P100G F144057M/P 10-100µL	Min. 10	± 0.35	≤ 0.10	3.5%	1%
	50	± 0.40	≤ 0.12	0.8%	0.24%
	Max. 100	± 0.80	≤ 0.15	0.8%	0.15%
P200G F144058M/P 20-200µL	Min. 20	± 0.50	≤ 0.20	2.5%	1%
	100	± 0.80	≤ 0.25	0.8%	0.25%
	Max. 200	± 1.60	≤ 0.30	0.8%	0.15%
P1000G F144059M/P 100-1000µL	Min. 100	± 3	≤ 0.6	3%	0.6%
	500	± 4	≤ 1.0	0.8%	0.2%
	Max. 1000	± 8	≤ 1.5	0.8%	0.15%
P5000G F144066 500-5000µL	Min. 500	± 12	≤ 3	2.4%	0.6%
	2500	± 15	≤ 5	0.6%	0.2%
	Max. 5000	± 30	≤ 8	0.6%	0.16%
P10mL G F144067 1-10mL	Min. 1 mL	± 30	≤ 6	3%	0.6%
	5 mL	± 40	≤ 10	0.8%	0.2%
	Max. 10 mL	± 60	≤ 16	0.6%	0.16%

## PIPETMAN Neo

		Pipette Specifications			
Model Part Number Volume Range	Volume (µL)	Systematic Error (µL)	Random Error (µL)	Systematic Error %	Random Error %
P2N F144561 0.2-2 µL	Min. 0.2	± 0.024	≤ 0.012	12%	6%
	0.5	± 0.025	≤ 0.012	5.0%	2.4%
	Max. 2	± 0.030	≤ 0.014	1.5%	0.7%
P10N F144562 1-10µL	Min. 1	± 0.025	≤ 0.012	2.5%	1.2%
	5	± 0.075	≤ 0.030	1.5%	0.6%
	Max. 10	± 0.100	≤ 0.040	1%	0.4%
P20N F144563 2-20µL	Min. 2	± 0.10	≤ 0.03	5%	1.5%
	10	± 0.10	≤ 0.05	1%	0.5%
	Max. 20	± 0.20	≤ 0.06	1%	0.3%
P100N F144564 10-100µL	Min. 10	± 0.35	≤ 0.10	3.5%	1%
	50	± 0.40	≤ 0.12	0.8%	0.24%
	Max. 100	± 0.80	≤ 0.15	0.8%	0.15%
P200N F144565 20-200µL	Min. 20	± 0.50	≤ 0.20	2.5%	1%
	100	± 0.80	≤ 0.25	0.8%	0.25%
	Max. 200	± 1.60	≤ 0.30	0.8%	0.15%
P1000N F144566 100-1000µL	Min. 100	± 3	≤ 0.6	3%	0.6%
	500	± 4	≤ 1.0	0.8%	0.2%
	Max. 1000	± 8	≤ 1.5	0.8%	0.15%

## PIPETMAN Neo Multichannel

		Pipette Specifications			
Model Part Number Volume Range	Volume (µL)	Systematic Error (µL)	Random Error (µL)	Systematic Error %	Random Error %
8x20 F14401 2-20 µL	Min. 2	± 0.10	≤ 0.08	5%	4%
	10	± 0.20	≤ 0.10	2%	1%
	Max. 20	± 0.40	≤ 0.20	2%	1%
12x20 F14402 2-20µL	Min. 2	± 0.10	≤ 0.08	5%	4%
	10	± 0.20	≤ 0.10	2%	1%
	Max. 20	± 0.40	≤ 0.20	2%	1%
8x200 F14403 20-200µL	Min. 20	± 0.50	≤ 0.25	2.5%	1.25%
	100	± 1.0	≤ 0.40	1%	0.4%
	Max. 200	± 2.0	≤ 0.60	1%	0.3%
12x200 F14404 20-200µL	Min. 20	± 0.5	≤ 0.25	2.5%	1.25%
	100	± 1.0	≤ 0.40	1%	0.4%
	Max. 200	± 2.0	≤ 0.60	1%	0.3%

PIPETMAN L					
Model Part Number	Volume ( $\mu\text{L}$ )	Systematic Error ( $\mu\text{L}$ )	Random Error ( $\mu\text{L}$ )	Systematic Error %	Random Error %
P2L FA10001M/P 0.2-2 $\mu\text{L}$	Min. 0.2	$\pm 0.024$	$\leq 0.012$	12%	6%
	0.5	$\pm 0.025$	$\leq 0.012$	5.0%	2.4%
	1	$\pm 0.027$	$\leq 0.013$	2.7%	1.3%
	Max. 2	$\pm 0.030$	$\leq 0.014$	1.5%	0.7%
P10L FA10002M/P 1-10 $\mu\text{L}$	Min. 1	$\pm 0.025$	$\leq 0.012$	2.5%	1.2%
	5	$\pm 0.075$	$\leq 0.030$	1.5%	0.6%
	Max. 10	$\pm 0.100$	$\leq 0.040$	1%	0.4%
P20L FA10003M/P 2-20 $\mu\text{L}$	Min. 2	$\pm 0.10$	$\leq 0.03$	5%	1.5%
	10	$\pm 0.10$	$\leq 0.05$	1%	0.5%
	Max. 20	$\pm 0.20$	$\leq 0.06$	1%	0.3%
P100L FA10004M/P 10-100 $\mu\text{L}$	Min. 10	$\pm 0.35$	$\leq 0.10$	3.5%	1%
	50	$\pm 0.40$	$\leq 0.12$	0.8%	0.24%
	Max. 100	$\pm 0.80$	$\leq 0.15$	0.8%	0.15%
P200L FA10005M/P 20-200 $\mu\text{L}$	Min. 20	$\pm 0.50$	$\leq 0.20$	2.5%	1%
	100	$\pm 0.80$	$\leq 0.25$	0.8%	0.25%
	Max. 200	$\pm 1.60$	$\leq 0.30$	0.8%	0.15%
P1000L FA10006M/P 100-1000 $\mu\text{L}$	Min. 100	$\pm 3$	$\leq 0.6$	3%	0.6%
	500	$\pm 4$	$\leq 1.0$	0.8%	0.2%
	Max. 1000	$\pm 8$	$\leq 1.5$	0.8%	0.15%
P5000L FA10007 500-5000 $\mu\text{L}$	Min. 500	$\pm 12$	$\leq 3$	2.4%	0.6%
	2500	$\pm 15$	$\leq 5$	0.6%	0.2%
	Max. 5000	$\pm 30$	$\leq 8$	0.6%	0.16%
P10mL L FA10008 1-10mL	Min. 1 mL	$\pm 30$	$\leq 6$	3%	0.6%
	5 mL	$\pm 40$	$\leq 10$	0.8%	0.2%
	Max. 10 mL	$\pm 60$	$\leq 16$	0.6%	0.16%

PIPETMAN M					
Model Part Number	Volume ( $\mu\text{L}$ )	Systematic Error ( $\mu\text{L}$ )	Random Error ( $\mu\text{L}$ )	Systematic Error %	Random Error %
P10M F81001 0.5-10 $\mu\text{L}$	Min. 0.5	$\pm 0.04$	$\leq 0.013$	8.0%	2.6%
	1	$\pm 0.025$	$\leq 0.012$	2.5%	1.2%
	5	$\pm 0.060$	$\leq 0.020$	1.2%	0.4%
	Max. 10	$\pm 0.080$	$\leq 0.025$	0.8%	0.25%
P20M F81002 2-20 $\mu\text{L}$	Min. 2	$\pm 0.075$	$\leq 0.025$	3.75%	1.25%
	10	$\pm 0.10$	$\leq 0.035$	1.0%	0.35%
	Max. 20	$\pm 0.15$	$\leq 0.05$	0.75%	0.25%
P200M F81003 20-200 $\mu\text{L}$	Min. 20	$\pm 0.40$	$\leq 0.15$	2.0%	0.75%
	100	$\pm 0.80$	$\leq 0.22$	0.8%	0.22%
	Max. 200	$\pm 1.0$	$\leq 0.26$	0.5%	0.13%
P1000M F81004 100-1000 $\mu\text{L}$	Min. 100	$\pm 2.5$	$\leq 0.4$	2.5%	0.4%
	500	$\pm 3.5$	$\leq 0.7$	0.7%	0.14%
	Max. 1000	$\pm 5.0$	$\leq 1.1$	0.5%	0.11%

PIPETMAN L Multichannel					
Model Part Number	Volume ( $\mu\text{L}$ )	Systematic Error ( $\mu\text{L}$ )	Random Error ( $\mu\text{L}$ )	Systematic Error %	Random Error %
P8x10L (F810013) P12x10L (F810014) 0.5-10 $\mu\text{L}$	Min. 0.5	$\pm 0.08$	$\leq 0.04$	16%	8.0%
	1	$\pm 0.08$	$\leq 0.05$	8%	5.0%
	5	$\pm 0.2$	$\leq 0.1$	4%	2%
	Max. 10	$\pm 0.2$	$\leq 0.1$	2%	1%
P8x20L (F810009) P12x20L (F810010) 2-20 $\mu\text{L}$	Min. 2	$\pm 0.10$	$\leq 0.08$	5.0%	4.0%
	10	$\pm 0.20$	$\leq 0.10$	2.0%	1.0%
	Max. 20	$\pm 0.40$	$\leq 0.15$	2.0%	0.75%
P8x200L (F810011) P12x200L (F810012) 20-200 $\mu\text{L}$	Min. 20	$\pm 0.50$	$\leq 0.25$	2.5%	1.25%
	100	$\pm 1.00$	$\leq 0.40$	1.0%	0.4%
	Max. 200	$\pm 2.00$	$\leq 0.50$	1.0%	0.25%
P8x300L (F810015) P12x300L (F810016) 20-300 $\mu\text{L}$	Min. 20	$\pm 1.0$	$\leq 0.35$	5.0%	1.75%
	30	$\pm 1.0$	$\leq 0.35$	3.3%	0.35%
	150	$\pm 1.5$	$\leq 0.60$	1.0%	0.4%
	Max. 300	$\pm 3.0$	$\leq 1.0$	1.0%	0.33%

PIPETMAN L - Fixed Volume					
Model Part Number	Volume ( $\mu\text{L}$ )	Systematic Error ( $\mu\text{L}$ )	Random Error ( $\mu\text{L}$ )	Systematic Error %	Random Error %
F1 - FA10017	1	$\pm 0.02$	$\leq 0.015$	2.0%	1.5%
F2 - FA10018	2	$\pm 0.05$	$\leq 0.02$	2.5%	1.0%
F5 - FA10019	5	$\pm 0.05$	$\leq 0.025$	1.0%	0.5%
F10 - FA10020	10	$\pm 0.06$	$\leq 0.03$	0.6%	0.3%
F20 - FA10021	20	$\pm 0.10$	$\leq 0.05$	0.5%	0.25%
F25 - FA10022	25	$\pm 0.20$	$\leq 0.07$	0.8%	0.28%
F50 - FA10023	50	$\pm 0.35$	$\leq 0.12$	0.7%	0.24%
F100 - FA10024	100	$\pm 0.55$	$\leq 0.15$	0.55%	0.15%
F200 - FA10025	200	$\pm 1.2$	$\leq 0.30$	0.6%	0.15%
F250 - FA10026	250	$\pm 1.5$	$\leq 0.75$	0.6%	0.3%
F300 - FA10027	300	$\pm 2.4$	$\leq 0.50$	0.8%	0.16%
F400 - FA10028	400	$\pm 2.4$	$\leq 0.80$	0.6%	0.2%
F500 - FA10029	500	$\pm 3.0$	$\leq 0.80$	0.6%	0.16%
F1000 - FA10030	1000	$\pm 5.0$	$\leq 1.3$	0.5%	0.13%
F5000 - FA10031	5000	$\pm 20.0$	$\leq 7.0$	0.4%	0.14%

### DISTRIMAN

Model Part Number Volume Range	Volume (µL)	Systematic Error (µL)	Random Error (µL)	Systematic Error %	Random Error %
<b>125µL Micro</b> F164100	Min. 2	± 0.100	≤ 0.080	5.0%	4.0%
	Max. 5	± 0.125	≤ 0.075	2.5%	1.5%
<b>125µL Micro ST*</b> F164130 1-12.5µL	Max. 10	± 0.200	≤ 0.100	2%	1.0%
<b>1250µL Mini</b> F164110	Min. 20	± 0.80	≤ 0.20	4.0%	1.0%
	Max. 50	± 1.0	≤ 0.40	2.0%	0.8%
<b>1250µL Mini ST*</b> F164140 10-125µL	Max. 100	± 1.0	≤ 0.60	1.0%	0.6%
<b>12.5mL Maxi</b> F164120	Min. 200	± 6.0	≤ 1.0	3.0%	0.5%
	Max. 500	± 7.5	≤ 1.5	1.5%	0.3%
<b>12.5mL Maxi ST</b> F164150 100-1.25mL	Max. 1000	± 10	≤ 2.5	1.0%	0.25%

### PIPETMAN Concept

Model Part Number Volume Range	Volume (µL)	Systematic Error (µL)	Random Error (µL)	Systematic Error %	Random Error %
<b>C10</b> F31012 0.5-10 µL	Min. 0.5	± 0.040	≤ 0.013	8.0%	2.6%
	1	± 0.025	≤ 0.012	2.5%	1.2%
	5	± 0.060	≤ 0.020	1.2%	0.4%
	Max. 10	± 0.080	≤ 0.025	0.8%	0.25%
<b>C100</b> F31013 5-100µL	Min. 5	± 0.35	≤ 0.10	7.0%	2.0%
	10	± 0.30	≤ 0.10	3.0%	1.0%
	50	± 0.38	≤ 0.12	0.76%	0.24%
	Max. 100	± 0.4	≤ 0.15	0.4%	0.15%
<b>C300</b> F31014 20-300µL	Min. 20	± 0.80	≤ 0.16	4.0%	0.8%
	30	± 0.70	≤ 0.20	2.33%	0.66%
	150	± 0.90	≤ 0.23	0.6%	0.15%
	Max. 300	± 1.05	≤ 0.30	0.35%	0.1%
<b>C1200</b> F31015 100-1200µL	Min. 100	± 2.5	≤ 0.4	2.5%	0.4%
	120	± 2.4	≤ 0.4	2.0%	0.33%
	600	± 3.6	≤ 0.8	0.6%	0.13%
	Max. 1200	± 6.0	≤ 1.2	0.5%	0.1%
<b>C5000</b> F31016 500-5000µL	Min. 500	± 10	≤ 2	2.0%	0.4%
	2500	± 15	≤ 4	0.6%	0.16%
	Max. 5000	± 25	≤ 7	0.5%	0.14%
<b>C10mL</b> F31017 1-10mL	Min. 1mL	± 25	≤ 4	2.5%	0.4%
	5mL	± 30	≤ 8	0.6%	0.16%
	Max. 10mL	± 50	≤ 12	0.5%	0.12%

### PIPETMAN Concept Multichannel

Model Part Number Volume Range	Volume (µL)	Systematic Error (µL)	Random Error (µL)	Systematic Error %	Random Error %
<b>C8x10</b> (F31032) <b>C12x10</b> (F31042) 0.5-10 µL	Min. 1	± 0.04	≤ 0.15	0.4%	0.15%
	5	± 0.08	≤ 0.04	1.6%	0.8%
	Max. 10	± 0.10	≤ 0.06	1.0%	0.6%
<b>C8x100</b> (F31033) <b>C12x100</b> (F31043) 5-100µL	Min. 10	± 0.25	≤ 0.14	2.5%	1.4%
	50	± 0.50	≤ 0.20	1.0%	0.4%
	Max. 100	± 0.80	≤ 0.25	0.8%	0.25%
<b>C8x300</b> (F31034) <b>C12x300</b> (F31044) 20-300µL	Min. 30	± 1.0	≤ 0.18	3.33%	0.6%
	150	± 1.5	≤ 0.38	1.0%	0.25%
	Max. 300	± 2.4	≤ 0.60	0.8	0.2%

### MICROMAN

Model Part Number Volume Range	Volume (µL)	Systematic Error (µL)	Random Error (µL)	Systematic Error %	Random Error %
<b>M10</b> F148501 1-10µL	Min. 1	± 0.09	≤ 0.03	9.0%	3.0%
	5	± 0.10	≤ 0.03	2.0%	0.6%
	Max. 10	± 0.15	≤ 0.06	1.5%	0.6%
<b>M25</b> F148502 3-25µL	Min. 3	± 0.25	≤ 0.08	8.33%	2.66%
	10	± 0.27	≤ 0.08	2.7%	0.8%
	Max. 25	± 0.30	≤ 0.10	1.2%	0.4%
<b>M50</b> F148503 20-50µL	Min. 20	± 0.34	≤ 0.20	1.7%	1.0%
	Max. 50	± 0.70	≤ 0.30	1.4%	0.6%
<b>M100</b> F148504 10-100 µL	Min. 10	± 0.50	≤ 0.20	5.0%	2.0%
	50	± 0.75	≤ 0.30	1.5%	0.6%
	Max. 100	± 1.0	≤ 0.40	1.0%	0.4%
<b>M250</b> F148505 50-250µL	Min. 50	± 1.5	≤ 0.30	3%	0.6%
	100	± 1.7	≤ 0.30	1.7%	0.3%
	Max. 250	± 2.5	≤ 0.50	1.0%	0.2%
<b>M1000</b> F148506 100-1000µL	Min. 100	± 3	≤ 1.6	3%	1.6%
	500	± 5	≤ 2.5	1%	0.5%
	Max. 1000	± 8	≤ 4	0.8%	0.4%