

Notified Body 0321

Issued to:

Globus (Shetland) Ltd Globus Building 38 Wheel Forge Way Trafford Park Manchester M17 1EH SATE

SATRA Client: P0130

EC Type-Examination Certificate

1572 Issue 5

This is to certify that the product group reference "SKYTEC UTAH" comprising the following products:

Product Ref.

Description

SKYTEC UTAH

Powder-free nitrile examination gloves in blue.

Classification			
EN 388:2003	Level	EN374-3:2003	Level
Abrasion resistance	0	Formaldehyde (15%)	6
Blade cut resistance	0	Sodium Hydroxide (40%)	6
Tear resistance	0	n-heptane	0
Puncture resistance	0	Sulphuric Acid (47%)	6

Technical reports:

SATRA: 26448/0151/SPC-0, 0155685/0736/BT,

TUV:7191004699-CHM11-01-TSL,7191004699-CHM11-02-TSL,

7191004699-CHM11-03-TSL.

has been subject to an EC Type-examination in accordance with Article 10 of the PPE Directive (89/686/EEC) and has been shown to satisfy the relevant provisions of this Directive for the complex category through:

- Testing to the following standard: EN 374-1:2003 excluding cause 5.3.2
- ii. Examination of the relevant technical documentation.

You are therefore licensed to mark the product(s) listed above in accordance with Article 13 of Directive (89/686/EEC) and any relevant amending Directives once you have drawn up an EC declaration of product conformity. Please note that:

- 1. Full details of the certification and product are contained in the manufacturer's technical file
- 2. This certificate is only valid if embossed with the text "SATRA European Notified Body 0321"
- 3. This certificate is issued subject to the conditions on the reverse side of this certificate
- 4. CE marking of production is also reliant on current compliance with Directive 89/686/EEC Article 11

Signed;

(D Cotter)

Signed:

Date: 10th September 2012

On behalf of SATRA